

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ENVI60
 Lab Code: ACE SDG NO.: Q2646
 Lab File ID: VN087368.D Date Analyzed: 07/21/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:38
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	202357	8.21	342303	9.08	307825	11.85
UPPER LIMIT	404714	8.706	684606	9.583	615650	12.347
LOWER LIMIT	101179	7.706	171152	8.583	153913	11.347
EPA SAMPLE NO.						
FRAC TANK	132035	8.21	262033	9.08	246052	11.85
VN0721WBL01	164239	8.21	328305	9.08	300772	11.85
VN0721WBS01	195078	8.21	337215	9.09	308523	11.85
VN0721WBSD01	190787	8.21	328848	9.08	296338	11.85

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.