

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2646</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Environmental Restoration, LLC</u>	Datafile :	<u>VN087372.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0721WBSD01	Dichlorodifluoromethane	20	22.4	ug/L	112	2		40 (69)	160 (116)	20 (19)
	Chloromethane	20	18.8	ug/L	94	0		40 (65)	160 (116)	20 (21)
	Vinyl chloride	20	19.0	ug/L	95	0		70 (65)	130 (117)	20 (19)
	Bromomethane	20	19.8	ug/L	99	3		40 (58)	160 (125)	20 (20)
	Chloroethane	20	18.5	ug/L	93	2		40 (56)	160 (128)	20 (20)
	Trichlorofluoromethane	20	19.7	ug/L	99	3		40 (73)	160 (115)	20 (16)
	1,1,2-Trichlorotrifluoroethane	20	19.6	ug/L	98	4		70 (80)	130 (112)	20 (15)
	1,1-Dichloroethene	20	17.6	ug/L	88	3		70 (74)	130 (110)	20 (20)
	Acetone	100	83.3	ug/L	83	4		40 (60)	160 (125)	20 (20)
	Carbon disulfide	20	18.6	ug/L	93	1		40 (64)	160 (112)	20 (20)
	Methyl tert-butyl Ether	20	18.0	ug/L	90	5		70 (78)	130 (114)	20 (20)
	Methyl Acetate	20	18.3	ug/L	92	4		70 (67)	130 (125)	20 (20)
	Methylene Chloride	20	17.7	ug/L	89	3		70 (72)	130 (114)	20 (20)
	trans-1,2-Dichloroethene	20	17.9	ug/L	90	1		70 (75)	130 (108)	20 (16)
	1,1-Dichloroethane	20	18.2	ug/L	91	3		70 (78)	130 (112)	20 (20)
	Cyclohexane	20	19.8	ug/L	99	2		70 (75)	130 (110)	20 (20)
	2-Butanone	100	88.2	ug/L	88	2		40 (65)	160 (122)	20 (26)
	Carbon Tetrachloride	20	19.8	ug/L	99	1		70 (77)	130 (113)	20 (15)
	cis-1,2-Dichloroethene	20	18.4	ug/L	92	2		70 (77)	130 (110)	20 (20)
	Bromochloromethane	20	18.6	ug/L	93	4		70 (70)	130 (124)	20 (20)
	Chloroform	20	18.4	ug/L	92	2		70 (79)	130 (113)	20 (20)
	1,1,1-Trichloroethane	20	18.9	ug/L	95	2		70 (80)	130 (108)	20 (20)
	Methylcyclohexane	20	20.9	ug/L	104	0		70 (72)	130 (115)	20 (20)
	Benzene	20	19.2	ug/L	96	1		70 (82)	130 (109)	20 (15)
	1,2-Dichloroethane	20	18.8	ug/L	94	2		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	18.8	ug/L	94	0		70 (77)	130 (113)	20 (15)
	1,2-Dichloropropane	20	19.3	ug/L	97	0		70 (83)	130 (111)	20 (16)
	Bromodichloromethane	20	18.8	ug/L	94	2		70 (83)	130 (110)	20 (16)
	4-Methyl-2-Pentanone	100	94.0	ug/L	94	2		40 (74)	160 (118)	20 (25)
	Toluene	20	19.5	ug/L	98	2		70 (82)	130 (110)	20 (16)
	t-1,3-Dichloropropene	20	18.6	ug/L	93	1		70 (79)	130 (110)	20 (20)
	cis-1,3-Dichloropropene	20	19.4	ug/L	97	3		70 (82)	130 (110)	20 (16)
	1,1,2-Trichloroethane	20	19.6	ug/L	98	4		70 (83)	130 (112)	20 (20)
	2-Hexanone	100	93.7	ug/L	94	2		40 (73)	160 (117)	20 (25)
	Dibromochloromethane	20	18.4	ug/L	92	2		70 (82)	130 (110)	20 (20)
	1,2-Dibromoethane	20	18.9	ug/L	95	9		70 (81)	130 (110)	20 (20)
	Tetrachloroethene	20	19.3	ug/L	97	5		70 (67)	130 (123)	20 (15)
	Chlorobenzene	20	18.6	ug/L	93	0		70 (82)	130 (109)	20 (15)
	Ethyl Benzene	20	18.7	ug/L	94	3		70 (83)	130 (109)	20 (16)
	m/p-Xylenes	40	38.8	ug/L	97	2		70 (82)	130 (110)	20 (15)
	o-Xylene	20	18.8	ug/L	94	2		70 (83)	130 (109)	20 (20)
	Styrene	20	19.4	ug/L	97	0		70 (80)	130 (111)	20 (17)
	Bromoform	20	18.8	ug/L	94	3		70 (79)	130 (109)	20 (20)
	Isopropylbenzene	20	19.3	ug/L	97	1		70 (83)	130 (112)	20 (29)
	1,1,2,2-Tetrachloroethane	20	19.0	ug/L	95	2		70 (76)	130 (118)	20 (20)
	1,3-Dichlorobenzene	20	19.3	ug/L	97	1		70 (82)	130 (108)	20 (20)
	1,4-Dichlorobenzene	20	18.4	ug/L	92	2		70 (82)	130 (107)	20 (15)
	1,2-Dichlorobenzene	20	19.7	ug/L	99	2		70 (82)	130 (109)	20 (20)

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2646</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Environmental Restoration, LLC</u>	Datafile :	<u>VN087372.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0721WBSD01	1,2-Dibromo-3-Chloropropane	20	16.6	ug/L	83	2		40 (68)	160 (112)	20 (20)
	1,2,4-Trichlorobenzene	20	18.5	ug/L	93	4		70 (75)	130 (113)	20 (29)
	1,2,3-Trichlorobenzene	20	18.4	ug/L	92	3		70 (76)	130 (114)	20 (29)