

## Initial and Continuing Calibration Verification

**Client:** Earth Engineering Inc.

**SDG No.:** Q2651

**Project:** Leon Avenue

**RunNo.:** LB136552

| Analyte               |      | Units | Result | True Value | %<br>Recovery | Acceptance<br>Window (%R) | Analysis<br>Date |
|-----------------------|------|-------|--------|------------|---------------|---------------------------|------------------|
| Sample ID:<br>Cyanide | ICV1 | mg/L  | 0.093  | 0.099      | 94            | 90-110                    | 07/21/2025       |
| Sample ID:<br>Cyanide | CCV1 | mg/L  | 0.25   | 0.25       | 100           | 90-110                    | 07/21/2025       |
| Sample ID:<br>Cyanide | CCV2 | mg/L  | 0.24   | 0.25       | 96            | 90-110                    | 07/21/2025       |
| Sample ID:<br>Cyanide | CCV3 | mg/L  | 0.24   | 0.25       | 96            | 90-110                    | 07/21/2025       |
| Sample ID:<br>Cyanide | CCV4 | mg/L  | 0.25   | 0.25       | 100           | 90-110                    | 07/21/2025       |