

Matrix Spike Summary

Client:	Earth Engineering Inc.	SDG No.:	Q2651
Project:	Leon Avenue	Sample ID:	Q2639-13
Client ID:	OU4-TS-44-071725MS	Percent Solids for Spike Sample:	64.9

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/Kg	75-125	2.80		0.069	J	3	1	91		07/21/2025

Matrix Spike Summary

Client:	Earth Engineering Inc.	SDG No.:	Q2651
Project:	Leon Avenue	Sample ID:	Q2639-13
Client ID:	OU4-TS-44-071725MSD	Percent Solids for Spike Sample:	64.9

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/Kg	75-125	2.90		0.069	J	3	1	94		07/21/2025