

Laboratory Control Sample Summary

Client: Earth Engineering Inc.

SDG No.: Q2651

Project: Leon Avenue

Run No.: LB136552

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168925BS							
Cyanide	mg/Kg	5	4.70		94	1	85-115	07/21/2025