

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Earth Engineering Inc.

SDG No.: Q2651

Contract: EARTH03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV34	Mercury	3.76	4.0	94	90 - 110	CV	07/21/2025	11:44	LB136551

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Earth Engineering Inc.

SDG No.: Q2651

Contract: EARTH03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV28	Mercury	4.91	5.0	98	90 - 110	CV	07/21/2025	11:49	LB136551
CCV29	Mercury	5.02	5.0	100	90 - 110	CV	07/21/2025	12:28	LB136551
CCV30	Mercury	5.18	5.0	104	90 - 110	CV	07/21/2025	12:56	LB136551
CCV31	Mercury	5.11	5.0	102	90 - 110	CV	07/21/2025	13:09	LB136551

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Earth Engineering Inc.

SDG No.: Q2651

Contract: EARTH03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	8040	8000	100	90 - 110	P	07/22/2025	12:03	LB136571
	Antimony	4100	4000	103	90 - 110	P	07/22/2025	12:03	LB136571
	Arsenic	3880	4000	97	90 - 110	P	07/22/2025	12:03	LB136571
	Barium	8140	8000	102	90 - 110	P	07/22/2025	12:03	LB136571
	Beryllium	203	200	102	90 - 110	P	07/22/2025	12:03	LB136571
	Cadmium	1950	2000	98	90 - 110	P	07/22/2025	12:03	LB136571
	Calcium	20300	20000	102	90 - 110	P	07/22/2025	12:03	LB136571
	Chromium	819	800	102	90 - 110	P	07/22/2025	12:03	LB136571
	Cobalt	2030	2000	102	90 - 110	P	07/22/2025	12:03	LB136571
	Copper	1030	1000	103	90 - 110	P	07/22/2025	12:03	LB136571
	Iron	4060	4000	102	90 - 110	P	07/22/2025	12:03	LB136571
	Lead	3800	4000	95	90 - 110	P	07/22/2025	12:03	LB136571
	Magnesium	20200	20000	101	90 - 110	P	07/22/2025	12:03	LB136571
	Manganese	2030	2000	102	90 - 110	P	07/22/2025	12:03	LB136571
	Nickel	2040	2000	102	90 - 110	P	07/22/2025	12:03	LB136571
	Potassium	19900	20000	99	90 - 110	P	07/22/2025	12:03	LB136571
	Selenium	3880	4000	97	90 - 110	P	07/22/2025	12:03	LB136571
	Silver	1040	1000	104	90 - 110	P	07/22/2025	12:03	LB136571
	Sodium	19600	20000	98	90 - 110	P	07/22/2025	12:03	LB136571
	Thallium	3770	4000	94	90 - 110	P	07/22/2025	12:03	LB136571
	Vanadium	2010	2000	100	90 - 110	P	07/22/2025	12:03	LB136571
	Zinc	2050	2000	102	90 - 110	P	07/22/2025	12:03	LB136571

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Earth Engineering Inc.

SDG No.: Q2651

Contract: EARTH03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	108	100	108	80 - 120	P	07/22/2025	12:10	LB136571
	Antimony	52.5	50.0	105	80 - 120	P	07/22/2025	12:10	LB136571
	Arsenic	21.8	20.0	109	80 - 120	P	07/22/2025	12:10	LB136571
	Barium	101	100	102	80 - 120	P	07/22/2025	12:10	LB136571
	Beryllium	6.30	6.0	105	80 - 120	P	07/22/2025	12:10	LB136571
	Cadmium	5.75	6.0	96	80 - 120	P	07/22/2025	12:10	LB136571
	Calcium	2130	2000	107	80 - 120	P	07/22/2025	12:10	LB136571
	Chromium	10.8	10.0	108	80 - 120	P	07/22/2025	12:10	LB136571
	Cobalt	29.8	30.0	99	80 - 120	P	07/22/2025	12:10	LB136571
	Copper	21.7	20.0	108	80 - 120	P	07/22/2025	12:10	LB136571
	Iron	107	100	107	80 - 120	P	07/22/2025	12:10	LB136571
	Lead	13.2	12.0	110	80 - 120	P	07/22/2025	12:10	LB136571
	Magnesium	2200	2000	110	80 - 120	P	07/22/2025	12:10	LB136571
	Manganese	22.0	20.0	110	80 - 120	P	07/22/2025	12:10	LB136571
	Nickel	40.9	40.0	102	80 - 120	P	07/22/2025	12:10	LB136571
	Potassium	2100	2000	105	80 - 120	P	07/22/2025	12:10	LB136571
	Selenium	21.7	20.0	109	80 - 120	P	07/22/2025	12:10	LB136571
	Silver	10.4	10.0	104	80 - 120	P	07/22/2025	12:10	LB136571
	Sodium	1840	2000	92	80 - 120	P	07/22/2025	12:10	LB136571
	Thallium	39.5	40.0	99	80 - 120	P	07/22/2025	12:10	LB136571
	Vanadium	40.4	40.0	101	80 - 120	P	07/22/2025	12:10	LB136571
	Zinc	42.6	40.0	106	80 - 120	P	07/22/2025	12:10	LB136571

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Earth Engineering Inc.

SDG No.: Q2651

Contract: EARTH03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	10000	10000	100	90 - 110	P	07/22/2025	13:00	LB136571
	Antimony	5050	5000	101	90 - 110	P	07/22/2025	13:00	LB136571
	Arsenic	5010	5000	100	90 - 110	P	07/22/2025	13:00	LB136571
	Barium	10100	10000	101	90 - 110	P	07/22/2025	13:00	LB136571
	Beryllium	247	250	99	90 - 110	P	07/22/2025	13:00	LB136571
	Cadmium	2470	2500	99	90 - 110	P	07/22/2025	13:00	LB136571
	Calcium	25500	25000	102	90 - 110	P	07/22/2025	13:00	LB136571
	Chromium	984	1000	98	90 - 110	P	07/22/2025	13:00	LB136571
	Cobalt	2480	2500	99	90 - 110	P	07/22/2025	13:00	LB136571
	Copper	1260	1250	101	90 - 110	P	07/22/2025	13:00	LB136571
	Iron	5030	5000	101	90 - 110	P	07/22/2025	13:00	LB136571
	Lead	4920	5000	98	90 - 110	P	07/22/2025	13:00	LB136571
	Magnesium	25000	25000	100	90 - 110	P	07/22/2025	13:00	LB136571
	Manganese	2530	2500	101	90 - 110	P	07/22/2025	13:00	LB136571
	Nickel	2490	2500	100	90 - 110	P	07/22/2025	13:00	LB136571
	Potassium	24900	25000	100	90 - 110	P	07/22/2025	13:00	LB136571
	Selenium	5030	5000	101	90 - 110	P	07/22/2025	13:00	LB136571
	Silver	1260	1250	100	90 - 110	P	07/22/2025	13:00	LB136571
	Sodium	24500	25000	98	90 - 110	P	07/22/2025	13:00	LB136571
	Thallium	5170	5000	103	90 - 110	P	07/22/2025	13:00	LB136571
	Vanadium	2500	2500	100	90 - 110	P	07/22/2025	13:00	LB136571
	Zinc	2530	2500	101	90 - 110	P	07/22/2025	13:00	LB136571
CCV02	Aluminum	10100	10000	101	90 - 110	P	07/22/2025	14:05	LB136571
	Antimony	5150	5000	103	90 - 110	P	07/22/2025	14:05	LB136571
	Arsenic	5090	5000	102	90 - 110	P	07/22/2025	14:05	LB136571
	Barium	10500	10000	105	90 - 110	P	07/22/2025	14:05	LB136571
	Beryllium	250	250	100	90 - 110	P	07/22/2025	14:05	LB136571
	Cadmium	2490	2500	100	90 - 110	P	07/22/2025	14:05	LB136571
	Calcium	26000	25000	104	90 - 110	P	07/22/2025	14:05	LB136571
	Chromium	976	1000	98	90 - 110	P	07/22/2025	14:05	LB136571
	Cobalt	2510	2500	100	90 - 110	P	07/22/2025	14:05	LB136571
	Copper	1280	1250	102	90 - 110	P	07/22/2025	14:05	LB136571
	Iron	5020	5000	100	90 - 110	P	07/22/2025	14:05	LB136571
	Lead	4970	5000	99	90 - 110	P	07/22/2025	14:05	LB136571

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Earth Engineering Inc.

SDG No.: Q2651

Contract: EARTH03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	24700	25000	99	90 - 110	P	07/22/2025	14:05	LB136571
	Manganese	2570	2500	103	90 - 110	P	07/22/2025	14:05	LB136571
	Nickel	2520	2500	101	90 - 110	P	07/22/2025	14:05	LB136571
	Potassium	25600	25000	102	90 - 110	P	07/22/2025	14:05	LB136571
	Selenium	5100	5000	102	90 - 110	P	07/22/2025	14:05	LB136571
	Silver	1280	1250	102	90 - 110	P	07/22/2025	14:05	LB136571
	Sodium	23100	25000	93	90 - 110	P	07/22/2025	14:05	LB136571
	Thallium	4810	5000	96	90 - 110	P	07/22/2025	14:05	LB136571
	Vanadium	2510	2500	100	90 - 110	P	07/22/2025	14:05	LB136571
	Zinc	2570	2500	103	90 - 110	P	07/22/2025	14:05	LB136571
CCV03	Aluminum	10500	10000	105	90 - 110	P	07/22/2025	15:20	LB136571
	Antimony	5250	5000	105	90 - 110	P	07/22/2025	15:20	LB136571
	Arsenic	5210	5000	104	90 - 110	P	07/22/2025	15:20	LB136571
	Barium	10600	10000	106	90 - 110	P	07/22/2025	15:20	LB136571
	Beryllium	262	250	105	90 - 110	P	07/22/2025	15:20	LB136571
	Cadmium	2560	2500	102	90 - 110	P	07/22/2025	15:20	LB136571
	Calcium	26800	25000	107	90 - 110	P	07/22/2025	15:20	LB136571
	Chromium	1010	1000	101	90 - 110	P	07/22/2025	15:20	LB136571
	Cobalt	2560	2500	103	90 - 110	P	07/22/2025	15:20	LB136571
	Copper	1310	1250	104	90 - 110	P	07/22/2025	15:20	LB136571
	Iron	5150	5000	103	90 - 110	P	07/22/2025	15:20	LB136571
	Lead	5110	5000	102	90 - 110	P	07/22/2025	15:20	LB136571
	Magnesium	25900	25000	104	90 - 110	P	07/22/2025	15:20	LB136571
	Manganese	2660	2500	107	90 - 110	P	07/22/2025	15:20	LB136571
	Nickel	2590	2500	103	90 - 110	P	07/22/2025	15:20	LB136571
	Potassium	25800	25000	103	90 - 110	P	07/22/2025	15:20	LB136571
	Selenium	5230	5000	104	90 - 110	P	07/22/2025	15:20	LB136571
	Silver	1310	1250	105	90 - 110	P	07/22/2025	15:20	LB136571
	Sodium	23600	25000	94	90 - 110	P	07/22/2025	15:20	LB136571
	Thallium	5070	5000	102	90 - 110	P	07/22/2025	15:20	LB136571
	Vanadium	2610	2500	104	90 - 110	P	07/22/2025	15:20	LB136571
	Zinc	2650	2500	106	90 - 110	P	07/22/2025	15:20	LB136571
CCV04	Aluminum	10500	10000	105	90 - 110	P	07/22/2025	16:13	LB136571
	Antimony	5280	5000	106	90 - 110	P	07/22/2025	16:13	LB136571

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Earth Engineering Inc.

SDG No.: Q2651

Contract: EARTH03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	5220	5000	104	90 - 110	P	07/22/2025	16:13	LB136571
	Barium	10800	10000	108	90 - 110	P	07/22/2025	16:13	LB136571
	Beryllium	257	250	103	90 - 110	P	07/22/2025	16:13	LB136571
	Cadmium	2560	2500	102	90 - 110	P	07/22/2025	16:13	LB136571
	Calcium	27000	25000	108	90 - 110	P	07/22/2025	16:13	LB136571
	Chromium	1010	1000	101	90 - 110	P	07/22/2025	16:13	LB136571
	Cobalt	2570	2500	103	90 - 110	P	07/22/2025	16:13	LB136571
	Copper	1310	1250	105	90 - 110	P	07/22/2025	16:13	LB136571
	Iron	5270	5000	105	90 - 110	P	07/22/2025	16:13	LB136571
	Lead	5100	5000	102	90 - 110	P	07/22/2025	16:13	LB136571
	Magnesium	25900	25000	104	90 - 110	P	07/22/2025	16:13	LB136571
	Manganese	2670	2500	107	90 - 110	P	07/22/2025	16:13	LB136571
	Nickel	2590	2500	104	90 - 110	P	07/22/2025	16:13	LB136571
	Potassium	26500	25000	106	90 - 110	P	07/22/2025	16:13	LB136571
	Selenium	5240	5000	105	90 - 110	P	07/22/2025	16:13	LB136571
	Silver	1320	1250	105	90 - 110	P	07/22/2025	16:13	LB136571
	Sodium	23800	25000	95	90 - 110	P	07/22/2025	16:13	LB136571
	Thallium	5150	5000	103	90 - 110	P	07/22/2025	16:13	LB136571
	Vanadium	2620	2500	105	90 - 110	P	07/22/2025	16:13	LB136571
	Zinc	2660	2500	106	90 - 110	P	07/22/2025	16:13	LB136571
CCV05	Aluminum	10400	10000	104	90 - 110	P	07/22/2025	17:14	LB136571
	Antimony	5210	5000	104	90 - 110	P	07/22/2025	17:14	LB136571
	Arsenic	5170	5000	104	90 - 110	P	07/22/2025	17:14	LB136571
	Barium	10700	10000	107	90 - 110	P	07/22/2025	17:14	LB136571
	Beryllium	263	250	105	90 - 110	P	07/22/2025	17:14	LB136571
	Cadmium	2560	2500	102	90 - 110	P	07/22/2025	17:14	LB136571
	Calcium	27100	25000	108	90 - 110	P	07/22/2025	17:14	LB136571
	Chromium	1020	1000	102	90 - 110	P	07/22/2025	17:14	LB136571
	Cobalt	2560	2500	103	90 - 110	P	07/22/2025	17:14	LB136571
	Copper	1300	1250	104	90 - 110	P	07/22/2025	17:14	LB136571
	Iron	5300	5000	106	90 - 110	P	07/22/2025	17:14	LB136571
	Lead	5120	5000	102	90 - 110	P	07/22/2025	17:14	LB136571
	Magnesium	26100	25000	104	90 - 110	P	07/22/2025	17:14	LB136571
	Manganese	2680	2500	107	90 - 110	P	07/22/2025	17:14	LB136571

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Earth Engineering Inc.

SDG No.: Q2651

Contract: EARTH03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2590	2500	103	90 - 110	P	07/22/2025	17:14	LB136571
	Potassium	26000	25000	104	90 - 110	P	07/22/2025	17:14	LB136571
	Selenium	5180	5000	104	90 - 110	P	07/22/2025	17:14	LB136571
	Silver	1320	1250	106	90 - 110	P	07/22/2025	17:14	LB136571
	Sodium	23800	25000	95	90 - 110	P	07/22/2025	17:14	LB136571
	Thallium	5000	5000	100	90 - 110	P	07/22/2025	17:14	LB136571
	Vanadium	2610	2500	104	90 - 110	P	07/22/2025	17:14	LB136571
	Zinc	2650	2500	106	90 - 110	P	07/22/2025	17:14	LB136571
CCV06	Aluminum	10500	10000	105	90 - 110	P	07/22/2025	18:05	LB136571
	Antimony	5240	5000	105	90 - 110	P	07/22/2025	18:05	LB136571
	Arsenic	5170	5000	103	90 - 110	P	07/22/2025	18:05	LB136571
	Barium	10700	10000	107	90 - 110	P	07/22/2025	18:05	LB136571
	Beryllium	264	250	106	90 - 110	P	07/22/2025	18:05	LB136571
	Cadmium	2560	2500	102	90 - 110	P	07/22/2025	18:05	LB136571
	Calcium	27100	25000	108	90 - 110	P	07/22/2025	18:05	LB136571
	Chromium	1040	1000	104	90 - 110	P	07/22/2025	18:05	LB136571
	Cobalt	2570	2500	103	90 - 110	P	07/22/2025	18:05	LB136571
	Copper	1300	1250	104	90 - 110	P	07/22/2025	18:05	LB136571
	Iron	5390	5000	108	90 - 110	P	07/22/2025	18:05	LB136571
	Lead	5120	5000	102	90 - 110	P	07/22/2025	18:05	LB136571
	Magnesium	26400	25000	106	90 - 110	P	07/22/2025	18:05	LB136571
	Manganese	2700	2500	108	90 - 110	P	07/22/2025	18:05	LB136571
	Nickel	2590	2500	104	90 - 110	P	07/22/2025	18:05	LB136571
	Potassium	26300	25000	105	90 - 110	P	07/22/2025	18:05	LB136571
	Selenium	5170	5000	103	90 - 110	P	07/22/2025	18:05	LB136571
	Silver	1330	1250	107	90 - 110	P	07/22/2025	18:05	LB136571
	Sodium	24600	25000	98	90 - 110	P	07/22/2025	18:05	LB136571
	Thallium	4640	5000	93	90 - 110	P	07/22/2025	18:05	LB136571
	Vanadium	2630	2500	105	90 - 110	P	07/22/2025	18:05	LB136571
	Zinc	2690	2500	108	90 - 110	P	07/22/2025	18:05	LB136571
CCV07	Aluminum	10400	10000	104	90 - 110	P	07/22/2025	18:54	LB136571
	Antimony	5250	5000	105	90 - 110	P	07/22/2025	18:54	LB136571
	Arsenic	5190	5000	104	90 - 110	P	07/22/2025	18:54	LB136571
	Barium	10600	10000	106	90 - 110	P	07/22/2025	18:54	LB136571

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Earth Engineering Inc.

SDG No.: Q2651

Contract: EARTH03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	267	250	107	90 - 110	P	07/22/2025	18:54	LB136571
	Cadmium	2570	2500	103	90 - 110	P	07/22/2025	18:54	LB136571
	Calcium	26900	25000	107	90 - 110	P	07/22/2025	18:54	LB136571
	Chromium	1010	1000	101	90 - 110	P	07/22/2025	18:54	LB136571
	Cobalt	2580	2500	103	90 - 110	P	07/22/2025	18:54	LB136571
	Copper	1310	1250	105	90 - 110	P	07/22/2025	18:54	LB136571
	Iron	5140	5000	103	90 - 110	P	07/22/2025	18:54	LB136571
	Lead	5130	5000	103	90 - 110	P	07/22/2025	18:54	LB136571
	Magnesium	26200	25000	105	90 - 110	P	07/22/2025	18:54	LB136571
	Manganese	2680	2500	107	90 - 110	P	07/22/2025	18:54	LB136571
	Nickel	2600	2500	104	90 - 110	P	07/22/2025	18:54	LB136571
	Potassium	25200	25000	101	90 - 110	P	07/22/2025	18:54	LB136571
	Selenium	5190	5000	104	90 - 110	P	07/22/2025	18:54	LB136571
	Silver	1320	1250	105	90 - 110	P	07/22/2025	18:54	LB136571
	Sodium	23300	25000	93	90 - 110	P	07/22/2025	18:54	LB136571
	Thallium	4570	5000	91	90 - 110	P	07/22/2025	18:54	LB136571
	Vanadium	2620	2500	105	90 - 110	P	07/22/2025	18:54	LB136571
	Zinc	2650	2500	106	90 - 110	P	07/22/2025	18:54	LB136571
CCV08	Aluminum	10600	10000	106	90 - 110	P	07/22/2025	19:19	LB136571
	Antimony	5340	5000	107	90 - 110	P	07/22/2025	19:19	LB136571
	Arsenic	5260	5000	105	90 - 110	P	07/22/2025	19:19	LB136571
	Barium	10800	10000	108	90 - 110	P	07/22/2025	19:19	LB136571
	Beryllium	270	250	108	90 - 110	P	07/22/2025	19:19	LB136571
	Cadmium	2580	2500	103	90 - 110	P	07/22/2025	19:19	LB136571
	Calcium	27200	25000	109	90 - 110	P	07/22/2025	19:19	LB136571
	Chromium	1050	1000	105	90 - 110	P	07/22/2025	19:19	LB136571
	Cobalt	2600	2500	104	90 - 110	P	07/22/2025	19:19	LB136571
	Copper	1320	1250	105	90 - 110	P	07/22/2025	19:19	LB136571
	Iron	5340	5000	107	90 - 110	P	07/22/2025	19:19	LB136571
	Lead	5150	5000	103	90 - 110	P	07/22/2025	19:19	LB136571
	Magnesium	26600	25000	106	90 - 110	P	07/22/2025	19:19	LB136571
	Manganese	2720	2500	109	90 - 110	P	07/22/2025	19:19	LB136571
	Nickel	2610	2500	104	90 - 110	P	07/22/2025	19:19	LB136571
	Potassium	25900	25000	104	90 - 110	P	07/22/2025	19:19	LB136571

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Earth Engineering Inc.

SDG No.: Q2651

Contract: EARTH03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	5290	5000	106	90 - 110	P	07/22/2025	19:19	LB136571
	Silver	1330	1250	107	90 - 110	P	07/22/2025	19:19	LB136571
	Sodium	24500	25000	98	90 - 110	P	07/22/2025	19:19	LB136571
	Thallium	5010	5000	100	90 - 110	P	07/22/2025	19:19	LB136571
	Vanadium	2660	2500	106	90 - 110	P	07/22/2025	19:19	LB136571
	Zinc	2710	2500	108	90 - 110	P	07/22/2025	19:19	LB136571