

Continuing Calibration Report for SequenceID : FE072225AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE054954.D

Aliphatic C9-C12	8247261.000	60.000	3.325	6.960	137454.350	136055.079	-1.028
Aliphatic C12-C16	5891259.000	40.000	6.961	10.412	147281.475	140568.128	-4.776
Aliphatic C16-C21	8721423.000	60.000	10.413	13.791	145357.050	144393.643	-0.667
Aliphatic C21-C28	11030578.000	80.000	13.792	17.461	137882.225	144651.717	4.680
Aliphatic C28-C40	15930506.000	120.000	17.462	22.488	132754.217	138647.492	4.251
Aliphatic EPH	49821027.000	360.000	3.325	22.488	138391.742	140720.791	1.655