

Continuing Calibration Report for SequenceID : FE072225AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE054974.D

Aliphatic C9-C12	8003485.000	60.000	3.325	6.960	133391.417	136055.079	1.958
Aliphatic C12-C16	5691816.000	40.000	6.961	10.412	142295.400	140568.128	-1.229
Aliphatic C16-C21	8499038.000	60.000	10.413	13.791	141650.633	144393.643	1.900
Aliphatic C21-C28	10820484.000	80.000	13.792	17.461	135256.050	144651.717	6.495
Aliphatic C28-C40	15503906.000	120.000	17.462	22.488	129199.217	138647.492	6.815
Aliphatic EPH	48518729.000	360.000	3.325	22.488	134774.247	140720.791	4.226