

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name:	<u>Alliance</u>	Contract:	<u>EARTH03</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2651</u>
Instrument ID:	<u>ECD_O</u>	Calibration Date(s):	<u>07/08/2025</u> <u>07/08/2025</u>
		Calibration Times:	<u>14:06</u> <u>22:16</u>

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 = <u>PO112088.D</u>	RT 750 = <u>PO112089.D</u>
RT 500 = <u>PO112090.D</u>	RT 250 = <u>PO112091.D</u>	RT 050 = <u>PO112092.D</u>

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Aroclor-1016-1	(1)	4.76	4.76	4.76	4.76	4.76	4.76	4.66	4.86
Aroclor-1016-2	(2)	4.78	4.78	4.78	4.78	4.78	4.78	4.68	4.88
Aroclor-1016-3	(3)	4.83	4.84	4.84	4.84	4.84	4.84	4.74	4.94
Aroclor-1016-4	(4)	4.95	4.96	4.96	4.96	4.96	4.96	4.86	5.06
Aroclor-1016-5	(5)	5.21	5.21	5.21	5.21	5.21	5.21	5.11	5.31
Aroclor-1260-1	(1)	6.25	6.25	6.25	6.25	6.25	6.25	6.15	6.35
Aroclor-1260-2	(2)	6.44	6.44	6.44	6.44	6.44	6.44	6.34	6.54
Aroclor-1260-3	(3)	6.81	6.81	6.81	6.81	6.81	6.81	6.71	6.91
Aroclor-1260-4	(4)	7.07	7.07	7.07	7.07	7.07	7.07	6.97	7.17
Aroclor-1260-5	(5)	7.31	7.31	7.31	7.31	7.31	7.31	7.21	7.41
Decachlorobiphenyl		8.70	8.70	8.70	8.70	8.70	8.70	8.60	8.80
Tetrachloro-m-xylene		3.67	3.67	3.67	3.67	3.67	3.67	3.57	3.77
Aroclor-1242-1	(1)	4.76	4.76	4.76	4.76	4.76	4.76	4.66	4.86
Aroclor-1242-2	(2)	4.78	4.78	4.78	4.78	4.78	4.78	4.68	4.88
Aroclor-1242-3	(3)	4.84	4.84	4.84	4.84	4.84	4.84	4.74	4.94
Aroclor-1242-4	(4)	4.96	4.96	4.96	4.96	4.96	4.96	4.86	5.06
Aroclor-1242-5	(5)	5.61	5.61	5.61	5.61	5.61	5.61	5.51	5.71
Decachlorobiphenyl		8.70	8.70	8.70	8.70	8.70	8.70	8.60	8.80
Tetrachloro-m-xylene		3.67	3.67	3.67	3.67	3.67	3.67	3.57	3.77
Aroclor-1248-1	(1)	4.76	4.76	4.76	4.76	4.76	4.76	4.66	4.86
Aroclor-1248-2	(2)	5.00	5.00	5.00	5.00	5.00	5.00	4.90	5.10
Aroclor-1248-3	(3)	5.21	5.21	5.21	5.21	5.21	5.21	5.11	5.31
Aroclor-1248-4	(4)	5.57	5.56	5.57	5.57	5.57	5.57	5.47	5.67
Aroclor-1248-5	(5)	5.61	5.61	5.61	5.61	5.61	5.61	5.51	5.71
Decachlorobiphenyl		8.70	8.70	8.70	8.70	8.70	8.70	8.60	8.80
Tetrachloro-m-xylene		3.67	3.67	3.67	3.67	3.67	3.67	3.57	3.77
Aroclor-1254-1	(1)	5.57	5.56	5.57	5.57	5.57	5.57	5.47	5.67
Aroclor-1254-2	(2)	5.72	5.71	5.72	5.71	5.71	5.72	5.62	5.82
Aroclor-1254-3	(3)	6.12	6.12	6.12	6.12	6.12	6.12	6.02	6.22
Aroclor-1254-4	(4)	6.35	6.35	6.35	6.35	6.35	6.35	6.25	6.45
Aroclor-1254-5	(5)	6.77	6.77	6.77	6.77	6.77	6.77	6.67	6.87
Decachlorobiphenyl		8.70	8.70	8.70	8.70	8.70	8.70	8.60	8.80
Tetrachloro-m-xylene		3.67	3.67	3.67	3.67	3.67	3.67	3.57	3.77
Aroclor-1268-1	(1)	7.59	7.59	7.59	7.59	7.59	7.59	7.49	7.69
Aroclor-1268-2	(2)	7.66	7.66	7.66	7.66	7.66	7.66	7.56	7.76
Aroclor-1268-3	(3)	7.86	7.86	7.86	7.86	7.86	7.86	7.76	7.96
Aroclor-1268-4	(4)	8.15	8.15	8.15	8.15	8.15	8.15	8.05	8.25
Aroclor-1268-5	(5)	8.45	8.44	8.45	8.45	8.45	8.45	8.35	8.55

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: EARTH03

Lab Code: ACE

SDG NO.: Q2651

Instrument ID: ECD_O

Calibration Date(s): **07/08/2025** **07/08/2025**

Calibration Times: 14:06 22:16

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 =	<u>PO112088.D</u>	RT 750 =	<u>PO112089.D</u>	
RT 500 =	PO112090.D	RT 250 =	PO112091.D	RT 050 =	PO112092.D

[illegible]

[illegible]