

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: EARTH03

Lab Code: ACE

SDG NO.: Q2651

Continuing Calib Date: 07/23/2025

Initial Calibration Date(s): 07/23/2025

07/23/2025

Continuing Calib Time: 13:22

Initial Calibration Time(s): 02:59

09:21

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.48	4.48	4.38	4.58	0.00
Aroclor-1248-2 (2)	4.75	4.75	4.65	4.85	0.00
Aroclor-1248-3 (3)	4.93	4.93	4.83	5.03	0.00
Aroclor-1248-4 (4)	5.32	5.32	5.22	5.42	0.00
Aroclor-1248-5 (5)	5.36	5.36	5.26	5.46	0.00
Tetrachloro-m-xylene	3.40	3.40	3.30	3.50	0.00
Decachlorobiphenyl	8.50	8.50	8.40	8.60	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>EARTH03</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2651</u>
Continuing Calib Date:	<u>07/23/2025</u>	Initial Calibration Date(s):	<u>07/23/2025</u> <u>07/23/2025</u>
Continuing Calib Time:	<u>13:22</u>	Initial Calibration Time(s):	<u>02:59</u> <u>09:21</u>

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	3.78	3.78	3.68	3.88	0.00
Aroclor-1248-2 (2)	4.01	4.01	3.91	4.11	0.00
Aroclor-1248-3 (3)	4.04	4.04	3.94	4.14	0.00
Aroclor-1248-4 (4)	4.20	4.20	4.10	4.30	0.00
Aroclor-1248-5 (5)	4.58	4.58	4.48	4.68	0.01
Tetrachloro-m-xylene	2.78	2.78	2.68	2.88	0.00
Decachlorobiphenyl	7.54	7.54	7.44	7.64	0.00