

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: EARTH03

Lab Code: ACE

SDG NO.: Q2651

Continuing Calib Date: 07/24/2025

Initial Calibration Date(s): 07/23/2025

07/23/2025

Continuing Calib Time: 05:15

Initial Calibration Time(s): 02:59

09:21

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.30	5.30	5.20	5.40	0.00
Aroclor-1254-2 (2)	5.51	5.51	5.41	5.61	0.00
Aroclor-1254-3 (3)	5.86	5.86	5.76	5.96	0.00
Aroclor-1254-4 (4)	6.14	6.14	6.04	6.24	0.00
Aroclor-1254-5 (5)	6.55	6.55	6.45	6.65	0.00
Tetrachloro-m-xylene	3.40	3.40	3.30	3.50	0.00
Decachlorobiphenyl	8.50	8.50	8.40	8.60	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: EARTH03

Lab Code: ACE

SDG NO.: Q2651

Continuing Calib Date: 07/24/2025

Initial Calibration Date(s): 07/23/2025

07/23/2025

Continuing Calib Time: 05:15

Initial Calibration Time(s): 02:59

09:21

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	4.54	4.54	4.44	4.64	0.00
Aroclor-1254-2 (2)	4.68	4.68	4.58	4.78	0.00
Aroclor-1254-3 (3)	5.07	5.07	4.97	5.17	0.00
Aroclor-1254-4 (4)	5.29	5.29	5.19	5.39	0.00
Aroclor-1254-5 (5)	5.69	5.70	5.60	5.80	0.01
Tetrachloro-m-xylene	2.78	2.78	2.68	2.88	0.00
Decachlorobiphenyl	7.54	7.54	7.44	7.64	0.00