

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** EARTH03
Lab Code: ACE **SDG NO.:** Q2651
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112366.D **Time Analyzed:** 08:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.759	4.661	4.861	518.970	500.000	3.8
Aroclor-1016-2	4.778	4.680	4.880	532.520	500.000	6.5
Aroclor-1016-3	4.835	4.736	4.936	514.470	500.000	2.9
Aroclor-1016-4	4.954	4.856	5.056	522.410	500.000	4.5
Aroclor-1016-5	5.211	5.113	5.313	511.450	500.000	2.3
Aroclor-1260-1	6.249	6.151	6.351	493.780	500.000	-1.2
Aroclor-1260-2	6.439	6.341	6.541	495.000	500.000	-1.0
Aroclor-1260-3	6.805	6.707	6.907	500.210	500.000	0.0
Aroclor-1260-4	7.064	6.967	7.167	503.510	500.000	0.7
Aroclor-1260-5	7.307	7.209	7.409	486.250	500.000	-2.8
Decachlorobiphenyl	8.700	8.603	8.803	49.980	50.000	0.0
Tetrachloro-m-xylene	3.672	3.573	3.773	57.200	50.000	14.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** EARTH03
Lab Code: ACE **SDG NO.:** Q2651
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112366.D **Time Analyzed:** 08:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.737	4.661	4.861	471.470	500.000	-5.7
Aroclor-1016-2	4.756	4.679	4.879	469.570	500.000	-6.1
Aroclor-1016-3	4.930	4.854	5.054	484.320	500.000	-3.1
Aroclor-1016-4	4.973	4.897	5.097	496.800	500.000	-0.6
Aroclor-1016-5	5.185	5.109	5.309	472.560	500.000	-5.5
Aroclor-1260-1	6.214	6.138	6.338	482.800	500.000	-3.4
Aroclor-1260-2	6.403	6.326	6.526	514.630	500.000	2.9
Aroclor-1260-3	6.554	6.478	6.678	517.520	500.000	3.5
Aroclor-1260-4	7.024	6.948	7.148	501.240	500.000	0.2
Aroclor-1260-5	7.267	7.190	7.390	519.750	500.000	4.0
Decachlorobiphenyl	8.643	8.568	8.768	47.160	50.000	-5.7
Tetrachloro-m-xylene	3.662	3.584	3.784	51.060	50.000	2.1