

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** EARTH03
Lab Code: ACE **SDG NO.:** Q2651
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ070611.D **Time Analyzed:** 22:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.483	4.383	4.583	540.470	500.000	8.1
Aroclor-1016-2	4.503	4.402	4.602	533.230	500.000	6.6
Aroclor-1016-3	4.560	4.460	4.660	531.100	500.000	6.2
Aroclor-1016-4	4.650	4.550	4.750	532.420	500.000	6.5
Aroclor-1016-5	4.934	4.834	5.034	532.180	500.000	6.4
Aroclor-1260-1	6.024	5.923	6.123	527.430	500.000	5.5
Aroclor-1260-2	6.280	6.180	6.380	527.990	500.000	5.6
Aroclor-1260-3	6.631	6.531	6.731	522.940	500.000	4.6
Aroclor-1260-4	6.844	6.745	6.945	525.760	500.000	5.2
Aroclor-1260-5	7.149	7.050	7.250	526.580	500.000	5.3
Decachlorobiphenyl	8.499	8.399	8.599	47.450	50.000	-5.1
Tetrachloro-m-xylene	3.403	3.303	3.503	52.260	50.000	4.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** EARTH03
Lab Code: ACE **SDG NO.:** Q2651
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ070611.D **Time Analyzed:** 22:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.781	3.681	3.881	481.780	500.000	-3.6
Aroclor-1016-2	3.797	3.697	3.897	478.940	500.000	-4.2
Aroclor-1016-3	3.957	3.857	4.057	476.990	500.000	-4.6
Aroclor-1016-4	4.006	3.906	4.106	468.790	500.000	-6.2
Aroclor-1016-5	4.201	4.101	4.301	486.800	500.000	-2.6
Aroclor-1260-1	5.194	5.094	5.294	472.270	500.000	-5.5
Aroclor-1260-2	5.383	5.284	5.484	471.060	500.000	-5.8
Aroclor-1260-3	5.524	5.424	5.624	465.510	500.000	-6.9
Aroclor-1260-4	5.984	5.885	6.085	455.660	500.000	-8.9
Aroclor-1260-5	6.229	6.129	6.329	447.410	500.000	-10.5
Decachlorobiphenyl	7.539	7.440	7.640	39.410	50.000	-21.2
Tetrachloro-m-xylene	2.781	2.680	2.880	46.240	50.000	-7.5