

Surrogate Summary

SDG No.: **Q2651**

Client: **Earth Engineering Inc.**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO112087.D	PIBLK-PO112087.D	Tetrachloro-m-xyl	1	20	19.4	97		60	140
		Decachlorobiphen	1	20	18.8	94		60	140
		Tetrachloro-m-xyl	2	20	20.0	100		60	140
		Decachlorobiphen	2	20	19.3	97		60	140
I.BLK-PO112370.D	PIBLK-PO112370.D	Tetrachloro-m-xyl	1	20	20.1	100		60	140
		Decachlorobiphen	1	20	17.7	89		60	140
		Tetrachloro-m-xyl	2	20	18.3	91		60	140
		Decachlorobiphen	2	20	17.4	87		60	140
PB168946BL	PB168946BL	Tetrachloro-m-xyl	1	20	20.9	104		32	144
		Decachlorobiphen	1	20	19.6	98		32	175
		Tetrachloro-m-xyl	2	20	18.2	91		32	144
		Decachlorobiphen	2	20	18.9	95		32	175
PB168946BS	PB168946BS	Tetrachloro-m-xyl	1	20	20.1	101		32	144
		Decachlorobiphen	1	20	20.3	102		32	175
		Tetrachloro-m-xyl	2	20	18.1	90		32	144
		Decachlorobiphen	2	20	19.5	98		32	175
Q2649-01MS	WC-1MS	Tetrachloro-m-xyl	1	20	24.5	122		32	144
		Decachlorobiphen	1	20	17.6	88		32	175
		Tetrachloro-m-xyl	2	20	22.2	111		32	144
		Decachlorobiphen	2	20	17.2	86		32	175
Q2649-01MSD	WC-1MSD	Tetrachloro-m-xyl	1	20	19.6	98		32	144
		Decachlorobiphen	1	20	17.8	89		32	175
		Tetrachloro-m-xyl	2	20	17.5	88		32	144
		Decachlorobiphen	2	20	17.4	87		32	175
I.BLK-PO112385.D	PIBLK-PO112385.D	Tetrachloro-m-xyl	1	20	19.8	99		60	140
		Decachlorobiphen	1	20	17.6	88		60	140
		Tetrachloro-m-xyl	2	20	18.1	90		60	140
		Decachlorobiphen	2	20	17.4	87		60	140
I.BLK-PQ070550.D	PIBLK-PQ070550.D	Tetrachloro-m-xyl	1	20	14.9	75		60	140
		Decachlorobiphen	1	20	17.0	85		60	140
		Tetrachloro-m-xyl	2	20	15.6	78		60	140
		Decachlorobiphen	2	20	17.5	88		60	140
I.BLK-PQ070600.D	PIBLK-PQ070600.D	Tetrachloro-m-xyl	1	20	20.5	102		60	140
		Decachlorobiphen	1	20	24.6	123		60	140
		Tetrachloro-m-xyl	2	20	19.8	99		60	140
		Decachlorobiphen	2	20	22.5	113		60	140
Q2651-01	MH 2-1	Tetrachloro-m-xyl	1	20	22.2	111		32	144
		Decachlorobiphen	1	20	21.6	108		32	175
		Tetrachloro-m-xyl	2	20	21.7	108		32	144
		Decachlorobiphen	2	20	17.7	88		32	175
Q2651-02	MH 6-5	Tetrachloro-m-xyl	1	20	19.7	98		32	144

Surrogate Summary

SDG No.: Q2651

Client: Earth Engineering Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q2651-02	MH 6-5	Decachlorobiphen	1	20	15.8	79		32	175
		Tetrachloro-m-xyl	2	20	19.1	95		32	144
Q2651-03	MH 7-6	Decachlorobiphen	2	20	12.0	60		32	175
		Tetrachloro-m-xyl	1	20	22.7	113		32	144
		Decachlorobiphen	1	20	19.7	98		32	175
		Tetrachloro-m-xyl	2	20	21.7	109		32	144
		Decachlorobiphen	2	20	15.4	77		32	175
Q2651-04	MH 8-7	Tetrachloro-m-xyl	1	20	20.7	103		32	144
		Decachlorobiphen	1	20	17.9	90		32	175
		Tetrachloro-m-xyl	2	20	19.8	99		32	144
		Decachlorobiphen	2	20	14.2	71		32	175
Q2651-05	MH 9-8	Tetrachloro-m-xyl	1	20	21.3	106		32	144
		Decachlorobiphen	1	20	16.8	84		32	175
		Tetrachloro-m-xyl	2	20	20.1	100		32	144
		Decachlorobiphen	2	20	12.2	61		32	175
I.BLK-PQ070615.D	PIBLK-PQ070615.D	Tetrachloro-m-xyl	1	20	20.0	100		60	140
		Decachlorobiphen	1	20	20.6	103		60	140
		Tetrachloro-m-xyl	2	20	18.5	93		60	140
		Decachlorobiphen	2	20	17.8	89		60	140