

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: EARTH03

Lab Code: ACE

SDG NO.: Q2651

Continuing Calib Date: 07/21/2025

Initial Calibration Date(s): 07/07/2025

07/07/2025

Continuing Calib Time: 21:12

Initial Calibration Time(s): 10:53

11:49

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.02	9.02	8.92	9.12	0.00
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
alpha-BHC	3.98	3.99	3.89	4.09	0.01
beta-BHC	4.50	4.50	4.40	4.60	0.00
delta-BHC	4.75	4.75	4.65	4.85	0.01
gamma-BHC (Lindane)	4.31	4.31	4.21	4.41	0.00
Heptachlor	4.91	4.91	4.81	5.01	0.00
Aldrin	5.25	5.25	5.15	5.35	0.01
Heptachlor epoxide	5.66	5.67	5.57	5.77	0.01
Endosulfan I	6.05	6.05	5.95	6.15	0.00
Dieldrin	6.32	6.32	6.22	6.42	0.00
4,4'-DDE	6.17	6.17	6.07	6.27	0.00
Endrin	6.55	6.55	6.45	6.65	0.00
Endosulfan II	6.76	6.76	6.66	6.86	0.00
4,4'-DDD	6.68	6.68	6.58	6.78	0.00
Endosulfan sulfate	7.12	7.12	7.02	7.22	0.00
4,4'-DDT	6.99	6.99	6.89	7.09	0.00
Methoxychlor	7.46	7.46	7.36	7.56	0.00
Endrin ketone	7.60	7.60	7.50	7.70	0.00
Endrin aldehyde	6.89	6.89	6.79	6.99	0.00
alpha-Chlordane	6.00	6.00	5.90	6.10	0.00
gamma-Chlordane	5.92	5.92	5.82	6.02	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: EARTH03

Lab Code: ACE

SDG NO.: Q2651

Continuing Calib Date: 07/21/2025

Initial Calibration Date(s): 07/07/2025

07/07/2025

Continuing Calib Time: 21:12

Initial Calibration Time(s): 10:53

11:49

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.99	8.00	7.90	8.10	0.01
Tetrachloro-m-xylene	2.83	2.83	2.73	2.93	0.00
alpha-BHC	3.33	3.34	3.24	3.44	0.01
beta-BHC	3.96	3.96	3.86	4.06	0.00
delta-BHC	4.20	4.20	4.10	4.30	0.00
gamma-BHC (Lindane)	3.67	3.67	3.57	3.77	0.00
Heptachlor	4.02	4.02	3.92	4.12	0.00
Aldrin	4.30	4.30	4.20	4.40	0.00
Heptachlor epoxide	4.80	4.80	4.70	4.90	0.00
Endosulfan I	5.17	5.17	5.07	5.27	0.00
Dieldrin	5.44	5.44	5.34	5.54	0.01
4,4'-DDE	5.31	5.31	5.21	5.41	0.01
Endrin	5.71	5.71	5.61	5.81	0.00
Endosulfan II	6.00	6.00	5.90	6.10	0.00
4,4'-DDD	5.86	5.86	5.76	5.96	0.00
Endosulfan sulfate	6.40	6.41	6.31	6.51	0.01
4,4'-DDT	6.11	6.11	6.01	6.21	0.00
Methoxychlor	6.68	6.68	6.58	6.78	0.00
Endrin ketone	6.91	6.91	6.81	7.01	0.00
Endrin aldehyde	6.18	6.18	6.08	6.28	0.00
alpha-Chlordane	5.12	5.12	5.02	5.22	0.00
gamma-Chlordane	5.05	5.05	4.95	5.15	0.00