

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072125\
 Data File : PL096493.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jul 2025 16:08
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 07/22/2025
 Supervised By :mohammad ahmed 07/23/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 22 01:47:08 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL070725.M
 Quant Title : GC Extractables
 QLast Update : Mon Jul 07 15:22:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.537	2.829	197.4E6	325.1E6	54.107	56.332
28)	SA Decachlor...	9.021	7.995	127.6E6	205.1E6	45.348	41.122
Target Compounds							
2)	A alpha-BHC	3.985	3.335	284.3E6	488.9E6	55.281	57.967
3)	MA gamma-BHC...	4.313	3.667	270.3E6	447.0E6	54.276	56.761
4)	MA Heptachlor	4.905	4.016	236.6E6	410.1E6	51.654	53.899
5)	MB Aldrin	5.244	4.299	248.5E6	408.6E6	51.409	56.009
6)	B beta-BHC	4.499	3.963	115.1E6	188.5E6	55.375	55.060
7)	B delta-BHC	4.745	4.196	251.7E6	434.2E6	56.170	56.375
8)	B Heptachlo...	5.663	4.801	220.8E6	356.3E6	53.207m	53.293
9)	A Endosulfan I	6.045	5.172	198.6E6	319.7E6	49.183m	44.558
10)	B gamma-Chl...	5.917	5.052	222.0E6	371.1E6	50.168m	52.871
11)	B alpha-Chl...	5.999	5.117	227.6E6	360.4E6	51.984	47.927
12)	B 4,4'-DDE	6.169	5.306	188.0E6	336.2E6	50.705	52.019
13)	MA Dieldrin	6.319	5.435	207.9E6	360.4E6	49.158	52.855m
14)	MA Endrin	6.545	5.711	159.3E6	317.2E6	51.234	52.033
15)	B Endosulfa...	6.756	6.003	171.9E6	287.0E6	49.546m	48.942
16)	A 4,4'-DDD	6.678	5.858	168.6E6	277.1E6	57.899	54.025
17)	MA 4,4'-DDT	6.992	6.111	140.3E6	242.6E6	46.217	42.301
18)	B Endrin al...	6.885	6.181	117.4E6	200.6E6	52.496m	45.024
19)	B Endosulfa...	7.120	6.405	147.2E6	270.0E6	48.111	48.012
20)	A Methoxychlor	7.464	6.683	74352715	123.2E6	47.196	40.572
21)	B Endrin ke...	7.599	6.909	156.5E6	288.8E6	47.399	46.189
22)	Mirex	8.080	7.100	127.2E6	212.1E6	45.727	42.694

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072125\
Data File : PL096493.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jul 2025 16:08
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/22/2025
Supervised By :mohammad ahmed 07/23/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 22 01:47:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL070725.M
Quant Title : GC Extractables
QLast Update : Mon Jul 07 15:22:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

