

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070725\
 Data File : PL096255.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Jul 2025 14:19
 Operator : AR\AJ
 Sample : PSTDICV050
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 ICVPL070725

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 07 15:23:32 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL070725.M
 Quant Title : GC Extractables
 QLast Update : Mon Jul 07 15:22:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.538	2.829	186.5E6	292.8E6	51.121	50.727
28)	SA Decachlor...	9.022	7.997	140.6E6	252.5E6	49.997	50.646
Target Compounds							
2)	A alpha-BHC	3.985	3.335	267.9E6	439.5E6	52.090	52.104
3)	MA gamma-BHC...	4.314	3.668	258.2E6	407.2E6	51.852	51.705
4)	MA Heptachlor	4.906	4.017	239.3E6	393.6E6	52.226	51.743
5)	MB Aldrin	5.246	4.300	251.0E6	377.0E6	51.928	51.670
6)	B beta-BHC	4.500	3.964	105.6E6	173.3E6	50.792	50.626
7)	B delta-BHC	4.746	4.197	234.7E6	398.9E6	52.366	51.789
8)	B Heptachlo...	5.665	4.802	246.7E6	343.2E6	59.453	51.344
9)	A Endosulfan I	6.047	5.174	205.2E6	331.1E6	50.831	46.148
10)	B gamma-Chl...	5.919	5.055	225.5E6	357.7E6	50.971	50.960
11)	B alpha-Chl...	6.000	5.119	222.4E6	359.0E6	50.784	47.738
12)	B 4,4'-DDE	6.170	5.308	190.6E6	334.4E6	51.411	51.739
13)	MA Dieldrin	6.320	5.438	215.9E6	351.2E6	51.070	51.494
14)	MA Endrin	6.546	5.713	165.8E6	314.2E6	53.309	51.541
15)	B Endosulfa...	6.758	6.005	180.6E6	300.7E6	52.046	51.281
16)	A 4,4'-DDD	6.679	5.860	150.5E6	262.9E6	51.690	51.256
17)	MA 4,4'-DDT	6.993	6.113	158.7E6	297.6E6	52.301	51.897
18)	B Endrin al...	6.887	6.183	122.0E6	220.2E6	54.555	49.434
19)	B Endosulfa...	7.120	6.406	156.8E6	287.8E6	51.259	51.179
20)	A Methoxychlor	7.465	6.685	80931127	156.5E6	51.372	51.507
21)	B Endrin ke...	7.600	6.911	170.6E6	319.2E6	51.674	51.060
22)	Mirex	8.080	7.102	139.8E6	249.2E6	50.233	50.167
24)	Chlordane-2	5.246f	0.000	251.0E6	0	1282.074	N.D. #
25)	Chlordane-3	5.919	5.055	225.5E6	357.7E6	306.122	444.629 #
26)	Chlordane-4	6.000	5.119	222.4E6	359.0E6	247.846	477.762 #
27)	Chlordane-5	0.000	6.005	0	300.7E6	N.D.	1060.880 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070725\
Data File : PL096255.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Jul 2025 14:19
Operator : AR\AJ
Sample : PSTDICV050
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
ICVPL070725

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 07 15:23:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL070725.M
Quant Title : GC Extractables
QLast Update : Mon Jul 07 15:22:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.538 min
Delta R.T.: 0.000 min
Response: 186539650
Conc: 51.12 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL070725




#1 Tetrachloro-m-xylene

R.T.: 2.829 min
Delta R.T.: 0.000 min
Response: 292757939
Conc: 50.73 ng/ml




#2 alpha-BHC

R.T.: 3.985 min
Delta R.T.: 0.001 min
Response: 267907777
Conc: 52.09 ng/ml



#2 alpha-BHC


R.T.: 3.335 min
Delta R.T.: 0.000 min
Response: 439467282
Conc: 52.10 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.314 min
Delta R.T.: 0.000 min
Response: 258204975
Conc: 51.85 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL070725




#3 gamma-BHC (Lindane)

R.T.: 3.668 min
Delta R.T.: 0.000 min
Response: 407224231
Conc: 51.71 ng/ml




#4 Heptachlor

R.T.: 4.906 min
Delta R.T.: 0.000 min
Response: 239257721
Conc: 52.23 ng/ml



#4 Heptachlor


R.T.: 4.017 min
Delta R.T.: 0.000 min
Response: 393645051
Conc: 51.74 ng/ml



#5 Aldrin


R.T.: 5.246 min
Delta R.T.: 0.000 min
Response: 251004503
Conc: 51.93 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL070725




#5 Aldrin

R.T.: 4.300 min
Delta R.T.: 0.000 min
Response: 376963069
Conc: 51.67 ng/ml




#6 beta-BHC

R.T.: 4.500 min
Delta R.T.: 0.000 min
Response: 105585547
Conc: 50.79 ng/ml



#6 beta-BHC


R.T.: 3.964 min
Delta R.T.: 0.000 min
Response: 173338446
Conc: 50.63 ng/ml



#7 delta-BHC


R.T.: 4.746 min
Delta R.T.: 0.000 min
Response: 234667452
Conc: 52.37 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL070725




#7 delta-BHC

R.T.: 4.197 min
Delta R.T.: 0.001 min
Response: 398872355
Conc: 51.79 ng/ml




#8 Heptachlor epoxide

R.T.: 5.665 min
Delta R.T.: 0.000 min
Response: 246681583
Conc: 59.45 ng/ml



#8 Heptachlor epoxide


R.T.: 4.802 min
Delta R.T.: 0.002 min
Response: 343231658
Conc: 51.34 ng/ml



#9 Endosulfan I


R.T.: 6.047 min
Delta R.T.: 0.000 min
Response: 205217213
Conc: 50.83 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL070725




#9 Endosulfan I

R.T.: 5.174 min
Delta R.T.: 0.002 min
Response: 331072403
Conc: 46.15 ng/ml




#10 gamma-Chlordane

R.T.: 5.919 min
Delta R.T.: 0.000 min
Response: 225536850
Conc: 50.97 ng/ml



#10 gamma-Chlordane


R.T.: 5.055 min
Delta R.T.: 0.000 min
Response: 357720812
Conc: 50.96 ng/ml



#11 alpha-Chlordane


R.T.: 6.000 min
Delta R.T.: 0.000 min
Response: 222374257
Conc: 50.78 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL070725




#11 alpha-Chlordane

R.T.: 5.119 min
Delta R.T.: 0.000 min
Response: 359003046
Conc: 47.74 ng/ml




#12 4,4'-DDE

R.T.: 6.170 min
Delta R.T.: 0.000 min
Response: 190603296
Conc: 51.41 ng/ml



#12 4,4'-DDE


R.T.: 5.308 min
Delta R.T.: 0.002 min
Response: 334423899
Conc: 51.74 ng/ml



#13 Dieldrin


R.T.: 6.320 min
Delta R.T.: 0.000 min
Response: 215943692
Conc: 51.07 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL070725




#13 Dieldrin

R.T.: 5.438 min
Delta R.T.: 0.002 min
Response: 351152654
Conc: 51.49 ng/ml




#14 Endrin

R.T.: 6.546 min
Delta R.T.: 0.000 min
Response: 165802977
Conc: 53.31 ng/ml



#14 Endrin


R.T.: 5.713 min
Delta R.T.: 0.002 min
Response: 314221022
Conc: 51.54 ng/ml



#15 Endosulfan II


R.T.: 6.758 min
Delta R.T.: 0.001 min
Response: 180601073
Conc: 52.05 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL070725




#15 Endosulfan II

R.T.: 6.005 min
Delta R.T.: 0.002 min
Response: 300706964
Conc: 51.28 ng/ml




#16 4,4'-DDD

R.T.: 6.679 min
Delta R.T.: 0.001 min
Response: 150504521
Conc: 51.69 ng/ml



#16 4,4'-DDD


R.T.: 5.860 min
Delta R.T.: 0.000 min
Response: 262873762
Conc: 51.26 ng/ml



#17 4,4'-DDT


R.T.: 6.993 min
Delta R.T.: 0.000 min
Response: 158743261
Conc: 52.30 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL070725




#17 4,4'-DDT

R.T.: 6.113 min
Delta R.T.: 0.000 min
Response: 297635340
Conc: 51.90 ng/ml




#18 Endrin aldehyde

R.T.: 6.887 min
Delta R.T.: 0.000 min
Response: 121960834
Conc: 54.56 ng/ml



#18 Endrin aldehyde


R.T.: 6.183 min
Delta R.T.: 0.002 min
Response: 220207892
Conc: 49.43 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.120 min
Delta R.T.: 0.000 min
Response: 156791775
Conc: 51.26 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL070725




#19 Endosulfan Sulfate

R.T.: 6.406 min
Delta R.T.: 0.000 min
Response: 287849227
Conc: 51.18 ng/ml




#20 Methoxychlor

R.T.: 7.465 min
Delta R.T.: 0.000 min
Response: 80931127
Conc: 51.37 ng/ml



#20 Methoxychlor


R.T.: 6.685 min
Delta R.T.: 0.000 min
Response: 156466302
Conc: 51.51 ng/ml



#21 Endrin ketone


R.T.: 7.600 min
Delta R.T.: 0.000 min
Response: 170609060
Conc: 51.67 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL070725




#21 Endrin ketone

R.T.: 6.911 min
Delta R.T.: 0.002 min
Response: 319242361
Conc: 51.06 ng/ml




#22 Mirex

R.T.: 8.080 min
Delta R.T.: 0.002 min
Response: 139758073
Conc: 50.23 ng/ml



#22 Mirex


R.T.: 7.102 min
Delta R.T.: 0.000 min
Response: 249239419
Conc: 50.17 ng/ml



#24 Chlordane-2


R.T.: 5.246 min
Delta R.T.: 0.029 min
Response: 251004503
Conc: 1282.07 ng/m

Instrument :
ECD_L
ClientSampleId :
ICVPL070725




#24 Chlordane-2

R.T.: 0.000 min
Exp R.T. : 4.420 min
Response: 0
Conc: N.D.




#25 Chlordane-3

R.T.: 5.919 min
Delta R.T.: 0.000 min
Response: 225536850
Conc: 306.12 ng/ml



#25 Chlordane-3


R.T.: 5.055 min
Delta R.T.: 0.000 min
Response: 357720812
Conc: 444.63 ng/ml



#26 Chlordane-4


R.T.: 6.000 min
Delta R.T.: -0.004 min
Response: 222374257
Conc: 247.85 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL070725




#26 Chlordane-4

R.T.: 5.119 min
Delta R.T.: 0.000 min
Response: 359003046
Conc: 477.76 ng/ml




#27 Chlordane-5

R.T.: 0.000 min
Exp R.T. : 6.841 min
Response: 0
Conc: N.D.



#27 Chlordane-5

R.T.: 6.005 min
Delta R.T.: -0.008 min
Response: 300706964
Conc: 1060.88 ng/ml



#28 Decachlorobiphenyl


R.T.: 9.022 min

Delta R.T.: 0.002 min

Response: 140644352

Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL070725



#28 Decachlorobiphenyl

R.T.: 7.997 min

Delta R.T.: 0.000 min

Response: 252544515

Conc: 50.65 ng/ml