

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2651

Client ID : SSTDCCC040

Date Analyzed: 07/21/2025

Lab File ID: BF143168.D

Time Analyzed: 13:26

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	129313	6.963	471327	8.25	224541	10.00
UPPER LIMIT	258626	7.463	942654	8.745	449082	10.498
LOWER LIMIT	64656.5	6.463	235664	7.745	112271	9.498
EPA SAMPLE NO.						
01 MH 2-1	127351	6.96	470127	8.24	227330	10.00
02 MH 6-5	136248	6.96	503155	8.24	241899	10.00
03 MH 7-6	124836	6.96	446383	8.24	207481	10.00
04 MH 8-7	141875	6.96	527265	8.24	244037	10.00
05 MH 9-8	134165	6.96	475638	8.24	208826	10.00

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.