

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: EARTH03
 Lab Code: ACE SDG NO.: Q2651
 Lab File ID: VY023015.D Date Analyzed: 07/21/2025
 Instrument ID: MSVOA_Y Time Analyzed: 10:08
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	315391	13.346				
UPPER LIMIT	630782	13.846				
LOWER LIMIT	157696	12.846				
EPA SAMPLE NO.						
MH 6-5	251674	13.34				
MH 7-6	250469	13.34				
MH 8-7	236661	13.34				
MH 9-8	6236 *	13.35				
VY0721SBL01	271493	13.35				
VY0721SBS01	283751	13.35				
VY0721SBSD01	278637	13.34				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.