

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: EARTH03
 Lab Code: ACE SDG NO.: Q2651
 Lab File ID: VY023039.D Date Analyzed: 07/22/2025
 Instrument ID: MSVOA_Y Time Analyzed: 11:04
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	265668	13.346				
UPPER LIMIT	531336	13.846				
LOWER LIMIT	132834	12.846				
EPA SAMPLE NO.						
MH 2-1	226724	13.35				
MH 9-8RE	28117 *	13.35				
VY0722SBL01	245159	13.35				
VY0722SBS01	248731	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.