

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q2653

Project: Parsippany Wastewater Quarterly 2025

RunNo.: LB136582

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.093	0.099	94	85-115	07/23/2025
Sample ID: Cyanide	CCV1	mg/L	0.25	0.25	100	90-110	07/23/2025
Sample ID: Cyanide	CCV2	mg/L	0.26	0.25	104	90-110	07/23/2025

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q2653

Project: Parsippany Wastewater Quarterly 2025

RunNo.: LB136589

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
Field pH		pH	7.03	7	100	90-110	07/21/2025
Sample ID:	CCV1						
Field pH		pH	7.01	7.0	100	90-110	07/21/2025
Sample ID:	CCV2						
Field pH		pH	7.01	7.0	100	90-110	07/21/2025