

Duplicate Sample Summary

Client: METEM A GE POWER Business	SDG No.: Q2653
Project: Parsippany Wastewater Quarterly 2025	Sample ID: Q2653-01
Client ID: WASTEWATER Q3DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.0012	U	0.0012	U	1	0		07/23/2025
Field pH	pH	+/-20	7.67		7.68		1	0.13		07/21/2025

Duplicate Sample Summary

Client: METEM A GE POWER Business	SDG No.: Q2653
Project: Parsippany Wastewater Quarterly 2025	Sample ID: Q2653-01
Client ID: WASTEWATER Q3MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.039		0.040		1	3		07/23/2025