

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: METEM A GE POWER Business

SDG No.: Q2653

Contract: METE01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Cadmium	2010	2000	100	95 - 105	P	07/23/2025	12:57	LB136588
	Chromium	783	800	98	95 - 105	P	07/23/2025	12:57	LB136588
	Copper	1010	1000	101	95 - 105	P	07/23/2025	12:57	LB136588
	Lead	3970	4000	99	95 - 105	P	07/23/2025	12:57	LB136588
	Nickel	1970	2000	98	95 - 105	P	07/23/2025	12:57	LB136588
	Silver	972	1000	97	95 - 105	P	07/23/2025	12:57	LB136588
	Zinc	2010	2000	100	95 - 105	P	07/23/2025	12:57	LB136588

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: METEM A GE POWER Business

SDG No.: Q2653

Contract: METE01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Cadmium	5.27	6.0	88	80 - 120	P	07/23/2025	13:02	LB136588
	Chromium	8.67	10.0	87	80 - 120	P	07/23/2025	13:02	LB136588
	Copper	21.5	20.0	108	80 - 120	P	07/23/2025	13:02	LB136588
	Lead	11.0	12.0	91	80 - 120	P	07/23/2025	13:02	LB136588
	Nickel	40.9	40.0	102	80 - 120	P	07/23/2025	13:02	LB136588
	Silver	10.6	10.0	106	80 - 120	P	07/23/2025	13:02	LB136588
	Zinc	43.0	40.0	108	80 - 120	P	07/23/2025	13:02	LB136588

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: METEM A GE POWER Business

SDG No.: Q2653

Contract: METE01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Cadmium	2470	2500	99	90 - 110	P	07/23/2025	13:42	LB136588
	Chromium	977	1000	98	90 - 110	P	07/23/2025	13:42	LB136588
	Copper	1270	1250	101	90 - 110	P	07/23/2025	13:42	LB136588
	Lead	4940	5000	99	90 - 110	P	07/23/2025	13:42	LB136588
	Nickel	2500	2500	100	90 - 110	P	07/23/2025	13:42	LB136588
	Silver	1270	1250	102	90 - 110	P	07/23/2025	13:42	LB136588
	Zinc	2550	2500	102	90 - 110	P	07/23/2025	13:42	LB136588
CCV02	Cadmium	2480	2500	99	90 - 110	P	07/23/2025	14:35	LB136588
	Chromium	973	1000	97	90 - 110	P	07/23/2025	14:35	LB136588
	Copper	1280	1250	102	90 - 110	P	07/23/2025	14:35	LB136588
	Lead	4940	5000	99	90 - 110	P	07/23/2025	14:35	LB136588
	Nickel	2500	2500	100	90 - 110	P	07/23/2025	14:35	LB136588
	Silver	1270	1250	102	90 - 110	P	07/23/2025	14:35	LB136588
	Zinc	2570	2500	103	90 - 110	P	07/23/2025	14:35	LB136588
CCV03	Cadmium	2400	2500	96	90 - 110	P	07/23/2025	15:57	LB136588
	Chromium	946	1000	95	90 - 110	P	07/23/2025	15:57	LB136588
	Copper	1240	1250	100	90 - 110	P	07/23/2025	15:57	LB136588
	Lead	4800	5000	96	90 - 110	P	07/23/2025	15:57	LB136588
	Nickel	2440	2500	98	90 - 110	P	07/23/2025	15:57	LB136588
	Silver	1240	1250	99	90 - 110	P	07/23/2025	15:57	LB136588
	Zinc	2520	2500	101	90 - 110	P	07/23/2025	15:57	LB136588
CCV04	Cadmium	2450	2500	98	90 - 110	P	07/23/2025	17:00	LB136588
	Chromium	975	1000	98	90 - 110	P	07/23/2025	17:00	LB136588
	Copper	1240	1250	99	90 - 110	P	07/23/2025	17:00	LB136588
	Lead	4860	5000	97	90 - 110	P	07/23/2025	17:00	LB136588
	Nickel	2470	2500	99	90 - 110	P	07/23/2025	17:00	LB136588
	Silver	1240	1250	99	90 - 110	P	07/23/2025	17:00	LB136588
	Zinc	2520	2500	101	90 - 110	P	07/23/2025	17:00	LB136588
CCV05	Cadmium	2480	2500	99	90 - 110	P	07/23/2025	17:49	LB136588
	Chromium	988	1000	99	90 - 110	P	07/23/2025	17:49	LB136588
	Copper	1260	1250	101	90 - 110	P	07/23/2025	17:49	LB136588
	Lead	4930	5000	98	90 - 110	P	07/23/2025	17:49	LB136588
	Nickel	2500	2500	100	90 - 110	P	07/23/2025	17:49	LB136588
	Silver	1270	1250	102	90 - 110	P	07/23/2025	17:49	LB136588

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: METEM A GE POWER Business

SDG No.: Q2653

Contract: METE01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Zinc	2570	2500	103	90 - 110	P	07/23/2025	17:49	LB136588
CCV06	Cadmium	2460	2500	98	90 - 110	P	07/23/2025	18:43	LB136588
	Chromium	979	1000	98	90 - 110	P	07/23/2025	18:43	LB136588
	Copper	1260	1250	101	90 - 110	P	07/23/2025	18:43	LB136588
	Lead	4880	5000	98	90 - 110	P	07/23/2025	18:43	LB136588
	Nickel	2480	2500	99	90 - 110	P	07/23/2025	18:43	LB136588
	Silver	1260	1250	101	90 - 110	P	07/23/2025	18:43	LB136588
	Zinc	2540	2500	102	90 - 110	P	07/23/2025	18:43	LB136588
CCV07	Cadmium	2420	2500	97	90 - 110	P	07/23/2025	19:38	LB136588
	Chromium	969	1000	97	90 - 110	P	07/23/2025	19:38	LB136588
	Copper	1240	1250	99	90 - 110	P	07/23/2025	19:38	LB136588
	Lead	4820	5000	96	90 - 110	P	07/23/2025	19:38	LB136588
	Nickel	2440	2500	98	90 - 110	P	07/23/2025	19:38	LB136588
	Silver	1240	1250	99	90 - 110	P	07/23/2025	19:38	LB136588
	Zinc	2510	2500	100	90 - 110	P	07/23/2025	19:38	LB136588
CCV08	Cadmium	2400	2500	96	90 - 110	P	07/23/2025	19:55	LB136588
	Chromium	963	1000	96	90 - 110	P	07/23/2025	19:55	LB136588
	Copper	1230	1250	98	90 - 110	P	07/23/2025	19:55	LB136588
	Lead	4810	5000	96	90 - 110	P	07/23/2025	19:55	LB136588
	Nickel	2430	2500	97	90 - 110	P	07/23/2025	19:55	LB136588
	Silver	1240	1250	99	90 - 110	P	07/23/2025	19:55	LB136588
	Zinc	2490	2500	100	90 - 110	P	07/23/2025	19:55	LB136588