



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: METEM A GE POWER Business

SDG No.: Q2653

Contract: METE01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Cadmium	5.67	6.0	94	65 - 135	P	07/23/2025	14:53	LB136588
	Chromium	9.46	10.0	95	65 - 135	P	07/23/2025	14:53	LB136588
	Copper	21.6	20.0	108	65 - 135	P	07/23/2025	14:53	LB136588
	Lead	11.3	12.0	94	65 - 135	P	07/23/2025	14:53	LB136588
	Nickel	39.9	40.0	100	65 - 135	P	07/23/2025	14:53	LB136588
	Silver	10.9	10.0	109	65 - 135	P	07/23/2025	14:53	LB136588
	Zinc	42.1	40.0	105	65 - 135	P	07/23/2025	14:53	LB136588