

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: METEM A GE POWER Business

SDG No.: Q2653

Contract: METE01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	-0.76	1.0	76	-5	7	07/23/2025	13:25	LB136588
	Chromium	56.9	52.0	109	42	62	07/23/2025	13:25	LB136588
	Copper	20.3	2.0	1015	-18	22	07/23/2025	13:25	LB136588
	Lead	-3.36			-12	12	07/23/2025	13:25	LB136588
	Nickel	3.84	2.0	192	-38	42	07/23/2025	13:25	LB136588
	Silver	1.88			-10	10	07/23/2025	13:25	LB136588
	Zinc	2.01			-40	40	07/23/2025	13:25	LB136588
ICSAB01	Cadmium	907	972	93	826	1120	07/23/2025	13:29	LB136588
	Chromium	528	542	97	460	624	07/23/2025	13:29	LB136588
	Copper	479	511	94	434	588	07/23/2025	13:29	LB136588
	Lead	39.7	49.0	81	37	61	07/23/2025	13:29	LB136588
	Nickel	952	954	100	810	1100	07/23/2025	13:29	LB136588
	Silver	208	201	104	170	232	07/23/2025	13:29	LB136588
	Zinc	1010	952	106	809	1095	07/23/2025	13:29	LB136588
ICSA	Cadmium	0.72	1.0	72	-5	7	07/23/2025	13:33	LB136588
	Chromium	48.5	52.0	93	42	62	07/23/2025	13:33	LB136588
	Copper	49.8	2.0	2490	-18	22	07/23/2025	13:33	LB136588
	Lead	-6.68			-12	12	07/23/2025	13:33	LB136588
	Nickel	10.9	2.0	543	-38	42	07/23/2025	13:33	LB136588
	Silver	11.6			-10	10	07/23/2025	13:33	LB136588
	Zinc	4.43			-40	40	07/23/2025	13:33	LB136588
ICSAB	Cadmium	876	972	90	826	1120	07/23/2025	13:37	LB136588
	Chromium	531	542	98	460	624	07/23/2025	13:37	LB136588
	Copper	499	511	98	434	588	07/23/2025	13:37	LB136588
	Lead	32.7	49.0	67	37	61	07/23/2025	13:37	LB136588
	Nickel	938	954	98	810	1100	07/23/2025	13:37	LB136588
	Silver	206	201	102	170	232	07/23/2025	13:37	LB136588
	Zinc	1010	952	106	809	1095	07/23/2025	13:37	LB136588