

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: METEM A GE POWER Business **Level:** LOW **SDG No.:** Q2653
Contract: METE01 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q2653-01 **Client ID:** WASTEWATER Q3DUP
Percent Solids for Sample: NA **Duplicate ID** Q2653-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	3.00	U	3.00	U			P
Chromium	ug/L	20	117		120		3		P
Copper	ug/L	20	179		182		2		P
Lead	ug/L	20	6.00	U	6.00	U			P
Nickel	ug/L	20	742		759		2		P
Silver	ug/L	20	1.44	J	1.68	J	15		P
Zinc	ug/L	20	48.8		50.1		3		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: METEM A GE POWER Business **Level:** LOW **SDG No.:** Q2653
Contract: METE01 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q2653-01MS **Client ID:** WASTEWATER Q3MSD
Percent Solids for Sample: NA **Duplicate ID** Q2653-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	86.7		81.9		6		P
Chromium	ug/L	20	296		281		5		P
Copper	ug/L	20	312		301		4		P
Lead	ug/L	20	416		394		5		P
Nickel	ug/L	20	958		930		3		P
Silver	ug/L	20	35.5		34.6		3		P
Zinc	ug/L	20	148		143		3		P