

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: METEM A GE POWER Business

SDG No.: Q2653

Contract: METE01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168955BS							
Cadmium	ug/L	100	94.3		94	85 - 115	P
Chromium	ug/L	200	199		100	85 - 115	P
Copper	ug/L	150	156		104	85 - 115	P
Lead	ug/L	500	472		94	85 - 115	P
Nickel	ug/L	250	250		100	85 - 115	P
Silver	ug/L	37.5	38.6		103	85 - 115	P
Zinc	ug/L	100	106		106	85 - 115	P