

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM070925

Review By	Rahul	Review On	7/9/2025 9:48:51 AM
Supervise By	Jagrut	Supervise On	7/9/2025 2:15:39 PM
SubDirectory	BM070925	HP Acquire Method	BNA_M
		HP Processing Method	bm070925

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12673,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM050376.D	08 Jul 2025 11:59		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BM050377.D	08 Jul 2025 12:39		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BM050378.D	08 Jul 2025 13:19		RC/JU	Ok
4	SSTDICC010	SSTDICC010	BM050379.D	08 Jul 2025 14:00		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BM050380.D	08 Jul 2025 14:40		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BM050381.D	08 Jul 2025 15:20	Compound#32,54,57,62,65 Kept on LR	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BM050382.D	08 Jul 2025 16:01		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BM050383.D	08 Jul 2025 16:41		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BM050384.D	08 Jul 2025 17:22	Compound#09 removed from 80 ppm	RC/JU	Ok
10	SSTDICV040	ICVBM070925	BM050385.D	08 Jul 2025 18:05	Comp#77 Failed from High side	RC/JU	Ok,M
11	PB168722BL	PB168722BL	BM050386.D	08 Jul 2025 19:26		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM072325

Review By	Rahul	Review On	7/24/2025 10:41:12 AM
Supervise By	Jagrut	Supervise On	7/24/2025 12:01:59 PM
SubDirectory	BM072325	HP Acquire Method	BNA_M
		HP Processing Method	bm070925

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S13168,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM050487.D	23 Jul 2025 12:35		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BM050488.D	23 Jul 2025 13:14		RC/JU	Ok
3	PB168971BL	PB168971BL	BM050489.D	23 Jul 2025 14:14		RC/JU	Ok
4	PB168971BS	PB168971BS	BM050490.D	23 Jul 2025 14:53		RC/JU	Ok,M
5	PB168971BSD	PB168971BSD	BM050491.D	23 Jul 2025 15:33	Injection error	RC/JU	Not Ok
6	Q2633-02	FIBER-GLASS-TANK	BM050492.D	23 Jul 2025 16:18		RC/JU	Ok
7	Q2664-01	GDW3	BM050493.D	23 Jul 2025 16:57		RC/JU	Ok
8	Q2668-04	TP-2	BM050494.D	23 Jul 2025 17:37	Analyzed with MSMSD, Not used	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP072125

Review By	Rahul	Review On	7/22/2025 8:59:18 AM
Supervise By	Jagrut	Supervise On	7/22/2025 9:10:39 AM
SubDirectory	BP072125	HP Acquire Method	BNA_P
		HP Processing Method	bp072125

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12674,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP025194.D	21 Jul 2025 13:02		CG/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP025195.D	21 Jul 2025 13:44	A Fresh Calibration is required.	CG/JU	Not Ok
3	SSTDICC2.5	SSTDICC2.5	BP025196.D	21 Jul 2025 14:25		CG/JU	Ok
4	SSTDICC005	SSTDICC005	BP025197.D	21 Jul 2025 15:06		CG/JU	Ok,M
5	SSTDICC010	SSTDICC010	BP025198.D	21 Jul 2025 15:47		CG/JU	Ok
6	SSTDICC020	SSTDICC020	BP025199.D	21 Jul 2025 16:29		CG/JU	Ok
7	SSTDICCC040	SSTDICCC040	BP025200.D	21 Jul 2025 17:10		CG/JU	Ok
8	SSTDICC050	SSTDICC050	BP025201.D	21 Jul 2025 17:51		CG/JU	Ok,M
9	SSTDICC060	SSTDICC060	BP025202.D	21 Jul 2025 18:32		CG/JU	Ok
10	SSTDICC080	SSTDICC080	BP025203.D	21 Jul 2025 19:14	Compound#79 removed from 80 PPM.	CG/JU	Ok,M
11	SSTDICV040	ICVBP072125	BP025204.D	21 Jul 2025 19:55		CG/JU	Ok,M
12	PB168854BL	PB168854BL	BP025205.D	21 Jul 2025 21:58		CG/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP072825

Review By	Rahul	Review On	7/28/2025 3:57:21 PM
Supervise By	Jagrut	Supervise On	7/29/2025 5:20:21 PM
SubDirectory	BP072825	HP Acquire Method	BNA_P
		HP Processing Method	bp072125

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S13168,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP025247.D	28 Jul 2025 11:59		CG/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP025248.D	28 Jul 2025 14:06		CG/JU	Ok,M
3	PB169022BL	PB169022BL	BP025249.D	28 Jul 2025 14:47		CG/JU	Ok
4	PB169022BS	PB169022BS	BP025250.D	28 Jul 2025 15:28		CG/JU	Ok,M
5	PB168971BL	PB168971BL	BP025251.D	28 Jul 2025 16:10		CG/JU	Ok
6	PB168971BS	PB168971BS	BP025252.D	28 Jul 2025 16:51		CG/JU	Ok,M
7	Q2703-01	TP-4	BP025253.D	28 Jul 2025 18:20	Internal Standard Fail	CG/JU	ReRun
8	PB168971BSD	PB168971BSD	BP025254.D	28 Jul 2025 19:02	.	CG/JU	Ok,M
9	Q2706-01	RT-5417	BP025255.D	28 Jul 2025 19:43		CG/JU	Ok
10	Q2706-03	ETGI-361	BP025256.D	28 Jul 2025 20:25		CG/JU	Ok
11	Q2706-03MS	ETGI-361MS	BP025257.D	28 Jul 2025 21:06		CG/JU	Ok
12	Q2706-03MSD	ETGI-361MSD	BP025258.D	28 Jul 2025 21:47		CG/JU	Ok,M
13	Q2700-01	EO-03-072525	BP025259.D	28 Jul 2025 22:28		CG/JU	Ok,M

M : Manual Integration