

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2664

Client ID : SSTDCCC040

Date Analyzed: 07/23/2025

Lab File ID: BM050488.D

Time Analyzed: 13:14

Instrument ID: BNA_M

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	538986	7.845	2095880	10.65	1417300	14.48
UPPER LIMIT	1077970	8.345	4191760	11.145	2834600	14.98
LOWER LIMIT	269493	7.345	1047940	10.145	708650	13.98
EPA SAMPLE NO.						
01 GDW3	754110	7.85	2702600	10.64	1716080	14.47

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.