

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2664

Client ID : SSTDCCC040

Date Analyzed: 07/28/2025

Lab File ID: BP025248.D

Time Analyzed: 14:06

Instrument ID: BNA\_P

GC Column: ZB-GR ID: 0.25 (mm)

|                | IS1 (DCB)<br>AREA # | RT #  | IS2 (NPT)<br>AREA # | RT #   | IS3 (ANT)<br>AREA # | RT #   |
|----------------|---------------------|-------|---------------------|--------|---------------------|--------|
| 12 HOUR STD    | 457120              | 7.414 | 1938820             | 10.15  | 1263470             | 14.04  |
| UPPER LIMIT    | 914240              | 7.914 | 3877640             | 10.649 | 2526940             | 14.543 |
| LOWER LIMIT    | 228560              | 6.914 | 969410              | 9.649  | 631735              | 13.543 |
| EPA SAMPLE NO. |                     |       |                     |        |                     |        |
| 01 PB168971BL  | 369641              | 7.42  | 1359800             | 10.16  | 821616              | 14.05  |
| 02 PB168971BS  | 370040              | 7.41  | 1420120             | 10.16  | 892276              | 14.05  |
| 03 PB168971BSD | 352130              | 7.43  | 1303880             | 10.16  | 807015              | 14.06  |

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.