

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2664

Client ID : SSTDCCC040

Date Analyzed: 07/30/2025

Lab File ID: BF143269.D

Time Analyzed: 15:22

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	110548	6.963	425847	8.25	214991	10.00
UPPER LIMIT	221096	7.463	851694	8.745	429982	10.504
LOWER LIMIT	55274	6.463	212924	7.745	107496	9.504
EPA SAMPLE NO.						
01 PB168971BL	102425	6.96	398799	8.24	212310	10.00
02 PB168971BS	110955	6.96	433977	8.25	229633	10.00
03 PB168971BSD	104468	6.96	406759	8.25	215723	10.00

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.