

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: GENV01  
 Lab Code: ACE SDG No.: Q2664  
 Instrument ID: MSVOA\_N Calibration Date/Time: 07/21/2025 10:38  
 Lab File ID: VN087368.D Init. Calib. Date(s): 07/16/2025 07/16/2025  
 Heated Purge: (Y/N) N Init. Calib. Time(s): 17:05 18:54  
 GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Methylene Chloride	50.000	50.57	-1.136002	20