

Surrogate Summary

SDG No.: **Q2664**

Client: **G Environmental**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2664-01	GDW3	1,2-Dichloroethane-d4	50	56.2	112		74	125
		Dibromofluoromethane	50	49.9	100		75	124
		Toluene-d8	50	51.5	103		86	113
		4-Bromofluorobenzene	50	50.5	101		77	121
VN0721WBL01	VN0721WBL01	1,2-Dichloroethane-d4	50	54.7	109		74	125
		Dibromofluoromethane	50	51.4	103		75	124
		Toluene-d8	50	51.0	102		86	113
		4-Bromofluorobenzene	50	46.4	93		77	121
VN0721WBS01	VN0721WBS01	1,2-Dichloroethane-d4	50	47.3	95		74	125
		Dibromofluoromethane	50	48.8	98		75	124
		Toluene-d8	50	50.4	101		86	113
		4-Bromofluorobenzene	50	49.9	100		77	121
VN0721WBSD0	VN0721WBSD01	1,2-Dichloroethane-d4	50	47.2	94		74	125
		Dibromofluoromethane	50	50.3	101		75	124
		Toluene-d8	50	50.6	101		86	113
		4-Bromofluorobenzene	50	50.0	100		77	121