

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q2667

Continuing Calib Date: 07/24/2025

Initial Calibration Date(s): 07/21/2025 07/21/2025

Continuing Calib Time: 12:03

Initial Calibration Time(s): 15:02 16:39

GC Column: RTX-CLP ID: 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|------------------|------------|-----------|------------------------|------|------------|
| 2,4-DCAA         | 7.32       | 7.33      | 7.23                   | 7.43 | 0.01       |
| 2,4-D            | 8.45       | 8.46      | 8.36                   | 8.56 | 0.01       |
| 2,4,5-TP(Silvex) | 9.34       | 9.34      | 9.24                   | 9.44 | 0.00       |

**CALIBRATION VERIFICATION SUMMARY**

**Lab Name:** Alliance

**Contract:** TETR16

**Lab Code:** ACE

**SDG NO.:** Q2667

**Continuing Calib Date:** 07/24/2025

**Initial Calibration Date(s):** 07/21/2025 07/21/2025

**Continuing Calib Time:** 12:03

**Initial Calibration Time(s):** 15:02 16:39

**GC Column:** RTX-CLP2 **ID:** 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|------------------|------------|-----------|-----------|-------|------------|
|                  |            |           | FROM      | TO    |            |
| 2,4-DCAA         | 7.76       | 7.77      | 7.67      | 7.87  | 0.01       |
| 2,4-D            | 9.02       | 9.02      | 8.92      | 9.12  | 0.00       |
| 2,4,5-TP(Silvex) | 9.92       | 9.93      | 9.83      | 10.03 | 0.01       |