

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q2667
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/24/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031243.D **Time Analyzed:** 23:29

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.335	9.242	9.442	685.930	712.500	-3.7
2,4-D	8.451	8.356	8.556	727.900	705.000	3.2
2,4-DCAA	7.320	7.225	7.425	669.120	750.000	-10.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q2667
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/24/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031243.D **Time Analyzed:** 23:29

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.926	9.829	10.029	666.400	712.500	-6.5
2,4-D	9.022	8.924	9.124	695.080	705.000	-1.4
2,4-DCAA	7.764	7.666	7.866	688.880	750.000	-8.1