

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2667</u>	Analytical Method:	<u>8151A</u>
Client:	<u>Tetra Tech, EMI</u>	DataFile :	<u>PS031235.D</u>

	Parameter	Sample Spike	Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q2641-02MSD (Column 1)	Client Sample ID:	P001-CONCRETE001-01M									
	2,4-D	50	0	76.2	ug/L	152	*	6		65	135	20
	2,4,5-TP(Silvex)	50	0	73.8	ug/L	148	*	5		62	139	20
Lab Sample ID:	Q2641-02MSD (Column 2)	Client Sample ID:	P001-CONCRETE001-01M									
	2,4-D	50	0	80.9	ug/L	162	*	6		65	135	20
	2,4,5-TP(Silvex)	50	0	72.7	ug/L	145	*	5		62	139	20