



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2667

Analytical Method: 8151A

Client: Tetra Tech, EMI

Datafile : PS031237.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB169001BS (Column 1)	2,4-D	5	6.00	ug/L	120				83	130	
	2,4,5-TP(Silvex)	5	5.60	ug/L	112				78	127	
PB169001BS (Column 2)	2,4-D	5	5.50	ug/L	110				83	130	
	2,4,5-TP(Silvex)	5	5.30	ug/L	106				78	127	