

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2669

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	1010	1000	101	95 - 105	P	07/23/2025	12:57	LB136588
	Iron	4000	4000	100	95 - 105	P	07/23/2025	12:57	LB136588
	Lead	3970	4000	99	95 - 105	P	07/23/2025	12:57	LB136588

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2669

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	21.5	20.0	108	80 - 120	P	07/23/2025	13:02	LB136588
	Iron	113	100	112	80 - 120	P	07/23/2025	13:02	LB136588
	Lead	11.0	12.0	91	80 - 120	P	07/23/2025	13:02	LB136588

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2669

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1270	1250	101	90 - 110	P	07/23/2025	13:42	LB136588
	Iron	4930	5000	99	90 - 110	P	07/23/2025	13:42	LB136588
	Lead	4940	5000	99	90 - 110	P	07/23/2025	13:42	LB136588
CCV02	Copper	1280	1250	102	90 - 110	P	07/23/2025	14:35	LB136588
	Iron	5060	5000	101	90 - 110	P	07/23/2025	14:35	LB136588
	Lead	4940	5000	99	90 - 110	P	07/23/2025	14:35	LB136588
CCV03	Copper	1240	1250	100	90 - 110	P	07/23/2025	15:57	LB136588
	Iron	4990	5000	100	90 - 110	P	07/23/2025	15:57	LB136588
	Lead	4800	5000	96	90 - 110	P	07/23/2025	15:57	LB136588
CCV04	Copper	1240	1250	99	90 - 110	P	07/23/2025	17:00	LB136588
	Iron	4890	5000	98	90 - 110	P	07/23/2025	17:00	LB136588
	Lead	4860	5000	97	90 - 110	P	07/23/2025	17:00	LB136588
CCV05	Copper	1260	1250	101	90 - 110	P	07/23/2025	17:49	LB136588
	Iron	4940	5000	99	90 - 110	P	07/23/2025	17:49	LB136588
	Lead	4930	5000	98	90 - 110	P	07/23/2025	17:49	LB136588
CCV06	Copper	1260	1250	101	90 - 110	P	07/23/2025	18:43	LB136588
	Iron	5070	5000	101	90 - 110	P	07/23/2025	18:43	LB136588
	Lead	4880	5000	98	90 - 110	P	07/23/2025	18:43	LB136588
CCV07	Copper	1240	1250	99	90 - 110	P	07/23/2025	19:38	LB136588
	Iron	4900	5000	98	90 - 110	P	07/23/2025	19:38	LB136588
	Lead	4820	5000	96	90 - 110	P	07/23/2025	19:38	LB136588
CCV08	Copper	1230	1250	98	90 - 110	P	07/23/2025	19:55	LB136588
	Iron	4890	5000	98	90 - 110	P	07/23/2025	19:55	LB136588
	Lead	4810	5000	96	90 - 110	P	07/23/2025	19:55	LB136588