

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client:	<u>Tully Environmental, Inc</u>	SDG No.:	<u>Q2669</u>
Contract:	<u>TULL01</u>	Lab Code:	<u>ACE</u>
ICS Source:	<u>EPA</u>	Instrument ID:	<u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	20.3	2.0	1015	-18	22	07/23/2025	13:25	LB136588
	Iron	102000	101000	101	85600	116500	07/23/2025	13:25	LB136588
	Lead	-3.36			-12	12	07/23/2025	13:25	LB136588
ICSAB01	Copper	479	511	94	434	588	07/23/2025	13:29	LB136588
	Iron	99600	99300	100	84400	114500	07/23/2025	13:29	LB136588
	Lead	39.7	49.0	81	37	61	07/23/2025	13:29	LB136588
ICSA	Copper	49.8	2.0	2490	-18	22	07/23/2025	13:33	LB136588
	Iron	97900	101000	97	85600	116500	07/23/2025	13:33	LB136588
	Lead	-6.68			-12	12	07/23/2025	13:33	LB136588
ICSAB	Copper	499	511	98	434	588	07/23/2025	13:37	LB136588
	Iron	97000	99300	98	84400	114500	07/23/2025	13:37	LB136588
	Lead	32.7	49.0	67	37	61	07/23/2025	13:37	LB136588