


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072225\
Data File : VN087388.D
Acq On : 22 Jul 2025 13:29
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N062525W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Wed Jun 25 10:49:56 2025



AutoFind: Averaged scan 1842 to 1856; Bkg corrected with scan 1841

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	23.1	1953	PASS
75	95	30	60	58.1	4917	PASS
95	95	100	100	100.0	8465	PASS
96	95	5	9	6.2	529	PASS
173	174	0.00	2	0.8	58	PASS
174	95	50	100	82.2	6954	PASS
175	174	5	9	7.0	490	PASS
176	174	95	101	96.4	6702	PASS
177	176	5	9	5.8	391	PASS