

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN062525

Review By	John Carlone	Review On	6/26/2025 8:30:28 AM
Supervise By	Maresh Dadoda	Supervise On	6/26/2025 2:41:09 PM
SubDirectory	VN062525	HP Acquire Method	HP Processing Method 624N062525W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP134497		
Initial Calibration Stds	VP134518,VP134519,VP134520,VP134521,VP134522		
CCC	VP134498,VP134524,VP134525		
Internal Standard/PEM			
ICV/I.BLK	VP134523		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN087161.D	25 Jun 2025 08:25	JC\MD	Ok
2	VSTDIC005	VN087162.D	25 Jun 2025 08:59	JC\MD	Ok,M
3	VSTDICCC020	VN087163.D	25 Jun 2025 09:20	JC\MD	Ok,M
4	VSTDIC050	VN087164.D	25 Jun 2025 09:41	JC\MD	Ok,M
5	VSTDIC0100	VN087165.D	25 Jun 2025 10:03	JC\MD	Ok,M
6	VSTDIC0150	VN087166.D	25 Jun 2025 10:24	JC\MD	Ok,M
7	IBLK	VN087167.D	25 Jun 2025 10:46	JC\MD	Ok
8	VSTDICV020	VN087168.D	25 Jun 2025 11:08	JC\MD	Ok,M
9	VN0625WBS01	VN087169.D	25 Jun 2025 11:41	JC\MD	Ok,M
10	VN0625WBSD01	VN087170.D	25 Jun 2025 12:15	JC\MD	Ok,M
11	VN0625WBL01	VN087171.D	25 Jun 2025 12:36	JC\MD	Ok
12	Q2349-01	VN087172.D	25 Jun 2025 13:11	JC\MD	Dilution
13	Q2349-04	VN087173.D	25 Jun 2025 13:33	JC\MD	Dilution
14	Q2349-01DL	VN087174.D	25 Jun 2025 13:54	JC\MD	Ok
15	Q2349-04DL	VN087175.D	25 Jun 2025 14:15	JC\MD	Ok
16	IBLK	VN087176.D	25 Jun 2025 14:37	JC\MD	Ok
17	Q2126-07 2.5PPB	VN087177.D	25 Jun 2025 14:58	JC\MD	Ok,M
18	Q2126-08 5.0PPB	VN087178.D	25 Jun 2025 15:20	JC\MD	Ok,M
19	VSTDCCC020	VN087179.D	25 Jun 2025 16:12	JC\MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN072225

Review By	Mahesh Dadoda	Review On	7/23/2025 8:17:54 AM
Supervise By	Semsettin Yesilyurt	Supervise On	7/23/2025 8:23:50 AM
SubDirectory	VN072225	HP Acquire Method	HP Processing Method 624N062525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134881		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134882,VP134883		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN087388.D	22 Jul 2025 13:29	JC\MD	Ok
2	VSTDCCC020	VN087389.D	22 Jul 2025 14:03	JC\MD	Ok,M
3	VN0722WBS01	VN087390.D	22 Jul 2025 15:14	JC\MD	Not Ok
4	VN0722WBS02	VN087391.D	22 Jul 2025 15:50	JC\MD	Ok,M
5	VN0722WBL01	VN087392.D	22 Jul 2025 16:12	JC\MD	Ok
6	Q2669-01	VN087393.D	22 Jul 2025 16:46	JC\MD	Dilution
7	Q2669-02	VN087394.D	22 Jul 2025 17:07	JC\MD	Dilution
8	Q2669-01DL	VN087395.D	22 Jul 2025 17:28	JC\MD	Ok,M
9	Q2669-02DL	VN087396.D	22 Jul 2025 17:49	JC\MD	Ok,M
10	VSTDCCC020	VN087397.D	22 Jul 2025 18:11	JC\MD	Ok,M

M : Manual Integration