

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TULL01
Lab Code: ACE SDG NO.: Q2669
Lab File ID: VN087389.D Date Analyzed: 07/22/2025
Instrument ID: MSVOA_N Time Analyzed: 14:03
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	26469	7.80	120974	9.08	115930	11.85
UPPER LIMIT	52938	8.3	241948	9.582	231860	12.347
LOWER LIMIT	13234.5	7.3	60487	8.582	57965	11.347
EPA SAMPLE NO.						
001 WILLETS PT BLVD (JUNE)	23863	7.80	105568	9.09	102924	11.85
001 WILLETS PT BLVD (JUNE)DL	14462	7.80	67398	9.08	62271	11.85
002 35th Ave (JUNE)	23359	7.79	108192	9.09	106953	11.85
002 35th Ave (JUNE)DL	25071	7.80	120236	9.08	108734	11.85
VN0722WBL01	27646	7.79	129340	9.09	109767	11.85
VN0722WBS02	28135	7.80	133851	9.09	128141	11.85

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.