

Report of Analysis

Client:	Tully Environmental, Inc		Date Collected:	07/21/25	
Project:	Transfer Station-SPDES		Date Received:	07/22/25	
Client Sample ID:	002 35th Ave (JUNE)		SDG No.:	Q2669	
Lab Sample ID:	Q2669-02		Matrix:	Water	
Analytical Method:	E624.1		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-BTEX	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087394.D	1	07/22/25 17:07	VN072225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
71-43-2	Benzene	0.45	U	0.45	5.00	ug/L
108-88-3	Toluene	820	E	0.46	5.00	ug/L
100-41-4	Ethyl Benzene	0.56	U	0.56	5.00	ug/L
179601-23-1	m/p-Xylenes	1.30	UQ	1.30	10.0	ug/L
95-47-6	o-Xylene	0.67	UQ	0.67	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	34.3	*	91 - 110	114%	SPK: 30
2037-26-5	Toluene-d8	27.6		91 - 112	92%	SPK: 30
460-00-4	4-Bromofluorobenzene	26.9		63 - 112	90%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	23400	7.794			
540-36-3	1,4-Difluorobenzene	108000	9.088			
3114-55-4	Chlorobenzene-d5	107000	11.847			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products