

DATA OF KNOWN QUALITY CONFORMANCE/NON-CONFORMANCE SUMMARY QUESTIONNAIRE

Laboratory Name : Alliance Technical Group LLC Client : RTP Environmental
 Project Location : _____ Project Number : _____
 Laboratory Sample ID(s) : Q2671 Sampling Date(s) : 7/22/2025
 List DKQP Methods Used (e.g., 8260,8270, et Cetra) **SMO,SOP,TO-15**

1	For each analytical method referenced in this laboratory report package, were all specified QA/QC performance criteria followed, including the requirement to explain any criteria falling outside of acceptable guidelines, as specified in the NJDEP Data of Known Quality performance standards?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
1A	Were the method specified handling, preservation, and holding time requirements met?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
1B	EPH Method: Was the EPH method conducted without significant modifications (see Section 11.3 of respective DKQ methods)	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A
2	Were all samples received by the laboratory in a condition consistent with that described on the associated chain-of-custody document(s)?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
3	Were samples received at an appropriate temperature (4±2° C)?	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A
4	Were all QA/QC performance criteria specified in the NJDEP DKQP standards achieved?	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No
5	a)Were reporting limits specified or referenced on the chain-of-custody or communicated to the laboratory prior to sample receipt? b)Were these reporting limits met?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
6	For each analytical method referenced in this laboratory report package, were results reported for all constituents identified in the method-specific analyte lists presented in the DKQP documents and/or site-specific QAPP?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
7	Are project-specific matrix spikes and/or laboratory duplicates included in this data set?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No

Notes: For all questions to which the response was "No" (with the exception of question #7), additional information should be provided in an attached narrative. If the answer to question #1, #1A, or #1B is "No", the data package does not meet the requirements for "Data of Known Quality."

Cover Page

Order ID : Q2671

Project ID : Yellow Cab - TO-15

Client : RTP Environmental

Lab Sample Number

Q2671-01
Q2671-02
Q2671-03
Q2671-04
Q2671-05

Client Sample Number

118-AAI
118-IA-1
118-IA-2
118-SG2
118-SG1

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 7/31/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

RTP Environmental

Project Name: Yellow Cab - TO-15

Project # N/A

Order ID # Q2671

Test Name: TO-15

A. Number of Samples and Date of Receipt:

5 Air samples were received on 07/22/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: TO-15. This data package contains results for TO-15.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_L were done using GC column RTX-1, which is 60 meters, 0.32 mm id, 1.0 um df, Restek Cat. #10157. The Trap was supplied by Entech, glass bead and Tenax , Entech 7100A Preconcentrator. The analysis of TO-15 was based on method TO-15.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Internal Standards Areas were met for all analysis.

The Retention Times were met for all analysis.

The RPD for {Q2671-01DUP} with File ID: VL042765.D met criteria except for Carbon Tetrachloride[28.6%] due to difference in results of original and DUP.

The Blank Spike met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The %RSD is greater than 30% in the Initial Calibration method (VL072325AIR.M) for Naphthalene and 1,2,4-Trichlorobenzene these compounds are passing on Linear Regression.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.

Due to potential high concentration of target analytes, Samples 118-SG1 was initially diluted.

Samples 118-IA-1, 118-IA-2, 118-SG2 and 118-SG1 were diluted due to high concentrations.

E. Additional Comments:

The Sample #118-AAI, 118-IA-1, 118-IA-2DL, 118-SG2, 118-SG1 have the concentration of target compound below Method detection limits, therefore it is not reported as Hit in Form1.

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

The Manual Integrations are performed for the followings.

Manual Integration Report			
Sequence	VL072325	Instrument	MSVOA_I

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICCO010	VL042747.D	m/p-Xylene	SAM	7/24/2025 9:02:56 AM	MMDado da	7/24/2025 2:20:57 PM	Peak Integrated by Software incorrectly
VSTDICCO02	VL042748.D	1,1,2-Trichloroethane	SAM	7/24/2025 9:03:01 AM	MMDado da	7/24/2025 2:20:59 PM	Peak Integrated by Software incorrectly
VSTDICCO02	VL042748.D	1,4-Dioxane	SAM	7/24/2025 9:03:01 AM	MMDado da	7/24/2025 2:20:59 PM	Peak Integrated by Software incorrectly
VSTDICCO02	VL042748.D	4-Methyl-2-Pentanone	SAM	7/24/2025 9:03:01 AM	MMDado da	7/24/2025 2:20:59 PM	Peak Integrated by Software incorrectly
VSTDICCO02	VL042748.D	cis-1,3-Dichloropropene	SAM	7/24/2025	MMDado da	7/24/2025	Peak Integrated

				9:03:01 AM		2:20:59 PM	ed by Software incorrectly
VSTDICCO 02	VL04274 8.D	Cyclohexane	SAM	7/24/20 25 9:03:01 AM	MMDado da	7/24/20 25 2:20:59 PM	Peak Integrated by Software incorrectly
VSTDICCO 02	VL04274 8.D	Heptane	SAM	7/24/20 25 9:03:01 AM	MMDado da	7/24/20 25 2:20:59 PM	Peak Integrated by Software incorrectly
VSTDICCO 02	VL04274 8.D	m/p-Xylene	SAM	7/24/20 25 9:03:01 AM	MMDado da	7/24/20 25 2:20:59 PM	Peak Integrated by Software incorrectly
VSTDICCO 02	VL04274 8.D	t-1,3-Dichloropropene	SAM	7/24/20 25 9:03:01 AM	MMDado da	7/24/20 25 2:20:59 PM	Peak Integrated by Software incorrectly
VSTDICCO 02	VL04274 8.D	trans-1,2-Dichloroethene	SAM	7/24/20 25 9:03:01 AM	MMDado da	7/24/20 25 2:20:59 PM	Peak Integrated by Software incorrectly
VSTDICCO 01	VL04274 9.D	1,4-Dioxane	SAM	7/24/20 25 9:03:21 AM	MMDado da	7/24/20 25 2:21:02 PM	Peak Integrated by Software incorrectly
VSTDICCO	VL04274	4-Methyl-2-	SAM	7/24/20	MMDado	7/24/20	Peak

01	9.D	Pentanone		25 9:03:21 AM	da	25 2:21:02 PM	Integrat ed by Softwar e incorrec tly
VSTDICCO 01	VL04274 9.D	Bromodichloromet hane	SAM	7/24/20 25 9:03:21 AM	MMDado da	7/24/20 25 2:21:02 PM	Peak Integrat ed by Softwar e incorrec tly
VSTDICCO 01	VL04274 9.D	cis-1,3- Dichloropropene	SAM	7/24/20 25 9:03:21 AM	MMDado da	7/24/20 25 2:21:02 PM	Peak Integrat ed by Softwar e incorrec tly
VSTDICCO 01	VL04274 9.D	Ethanol	SAM	7/24/20 25 9:03:21 AM	MMDado da	7/24/20 25 2:21:02 PM	Peak Integrat ed by Softwar e incorrec tly
VSTDICCO 01	VL04274 9.D	Heptane	SAM	7/24/20 25 9:03:21 AM	MMDado da	7/24/20 25 2:21:02 PM	Peak Integrat ed by Softwar e incorrec tly
VSTDICCO 01	VL04274 9.D	m/p-Xylene	SAM	7/24/20 25 9:03:21 AM	MMDado da	7/24/20 25 2:21:02 PM	Peak Integrat ed by Softwar e incorrec tly
VSTDICCO 01	VL04274 9.D	Methyl Methacrylate	SAM	7/24/20 25 9:03:21 AM	MMDado da	7/24/20 25 2:21:02 PM	Peak Integrat ed by Softwar e incorrec tly

VSTDICCO 01	VL04274 9.D	t-1,3- Dichloropropene	SAM	7/24/20 25 9:03:21 AM	MMDado da	7/24/20 25 2:21:02 PM	Peak Integrat ed by Softwar e incorrec tly
VSTDICCO. 5	VL04275 0.D	1,1,2- Trichloroethane	SAM	7/24/20 25 9:03:31 AM	MMDado da	7/24/20 25 2:21:04 PM	Peak Integrat ed by Softwar e incorrec tly
VSTDICCO. 5	VL04275 0.D	1,2- Dibromoethane	SAM	7/24/20 25 9:03:31 AM	MMDado da	7/24/20 25 2:21:04 PM	Peak Integrat ed by Softwar e incorrec tly
VSTDICCO. 5	VL04275 0.D	1,4-Dioxane	SAM	7/24/20 25 9:03:31 AM	MMDado da	7/24/20 25 2:21:04 PM	Peak Integrat ed by Softwar e incorrec tly
VSTDICCO. 5	VL04275 0.D	2-Hexanone	SAM	7/24/20 25 9:03:31 AM	MMDado da	7/24/20 25 2:21:04 PM	Peak Integrat ed by Softwar e incorrec tly
VSTDICCO. 5	VL04275 0.D	4-Methyl-2- Pentanone	SAM	7/24/20 25 9:03:31 AM	MMDado da	7/24/20 25 2:21:04 PM	Peak Integrat ed by Softwar e incorrec tly
VSTDICCO. 5	VL04275 0.D	Benzyl Chloride	SAM	7/24/20 25 9:03:31 AM	MMDado da	7/24/20 25 2:21:04 PM	Peak Integrat ed by Softwar e incorrec tly

VSTDICCO.5	VL04275 O.D	Bromodichloromethane	SAM	7/24/2025 9:03:31 AM	MMDado da	7/24/2025 2:21:04 PM	Peak Integrated by Software incorrectly
VSTDICCO.5	VL04275 O.D	cis-1,3-Dichloropropene	SAM	7/24/2025 9:03:31 AM	MMDado da	7/24/2025 2:21:04 PM	Peak Integrated by Software incorrectly
VSTDICCO.5	VL04275 O.D	Dibromochloromethane	SAM	7/24/2025 9:03:31 AM	MMDado da	7/24/2025 2:21:04 PM	Peak Integrated by Software incorrectly
VSTDICCO.5	VL04275 O.D	Ethanol	SAM	7/24/2025 9:03:31 AM	MMDado da	7/24/2025 2:21:04 PM	Peak Integrated by Software incorrectly
VSTDICCO.5	VL04275 O.D	Heptane	SAM	7/24/2025 9:03:31 AM	MMDado da	7/24/2025 2:21:04 PM	Peak Integrated by Software incorrectly
VSTDICCO.5	VL04275 O.D	m/p-Xylene	SAM	7/24/2025 9:03:31 AM	MMDado da	7/24/2025 2:21:04 PM	Peak Integrated by Software incorrectly
VSTDICCO.5	VL04275 O.D	Methyl Methacrylate	SAM	7/24/2025 9:03:31 AM	MMDado da	7/24/2025 2:21:04 PM	Peak Integrated by Software incorrectly

VSTDICCO.5	VL042750.D	t-1,3-Dichloropropene	SAM	7/24/2025 9:03:31 AM	MMDado da	7/24/2025 2:21:04 PM	Peak Integrated by Software incorrectly
VSTDICCO.5	VL042750.D	Tetrahydrofuran	SAM	7/24/2025 9:03:31 AM	MMDado da	7/24/2025 2:21:04 PM	Peak Integrated by Software incorrectly
VSTDICCO.5	VL042750.D	Trichloroethene	SAM	7/24/2025 9:03:31 AM	MMDado da	7/24/2025 2:21:04 PM	Peak Integrated by Software incorrectly
VSTDICCO.03	VL042752.D	1,1,1-Trichloroethane	SAM	7/24/2025 9:03:05 AM	MMDado da	7/24/2025 2:21:09 PM	Peak Integrated by Software incorrectly
VSTDICCO.03	VL042752.D	1,1,2,2-Tetrachloroethane	SAM	7/24/2025 9:03:05 AM	MMDado da	7/24/2025 2:21:09 PM	Peak Integrated by Software incorrectly
VSTDICCO.03	VL042752.D	Carbon Tetrachloride	SAM	7/24/2025 9:03:05 AM	MMDado da	7/24/2025 2:21:09 PM	Peak Integrated by Software incorrectly
VSTDICCO.03	VL042752.D	Tetrachloroethene	SAM	7/24/2025 9:03:05 AM	MMDado da	7/24/2025 2:21:09 PM	Peak Integrated by Software incorrectly

VSTDICCO.03	VL04275 2.D	Trichloroethene	SAM	7/24/20 25 9:03:05 AM	MMDado da	7/24/20 25 2:21:09 PM	Peak Integrated by Software incorrectly
VSTDICCO 15	VL04275 3.D	m/p-Xylene	SAM	7/24/20 25 9:03:41 AM	MMDado da	7/24/20 25 2:21:10 PM	Peak Integrated by Software incorrectly
VSTDICCO.1	VL04275 4.D	1,1,2,2-Tetrachloroethane	SAM	7/24/20 25 9:03:10 AM	MMDado da	7/24/20 25 2:21:12 PM	Peak Integrated by Software incorrectly
VSTDICCO.1	VL04275 4.D	1,2-Dibromoethane	SAM	7/24/20 25 9:03:10 AM	MMDado da	7/24/20 25 2:21:12 PM	Peak Integrated by Software incorrectly
VSTDICCO.1	VL04275 4.D	Carbon Tetrachloride	SAM	7/24/20 25 9:03:10 AM	MMDado da	7/24/20 25 2:21:12 PM	Peak Integrated by Software incorrectly
VSTDICCO.1	VL04275 4.D	Tetrachloroethene	SAM	7/24/20 25 9:03:10 AM	MMDado da	7/24/20 25 2:21:12 PM	Peak Integrated by Software incorrectly
VSTDICCO.1	VL04275 4.D	Trichloroethene	SAM	7/24/20 25 9:03:10 AM	MMDado da	7/24/20 25 2:21:12 PM	Peak Integrated by Software incorrectly

VSTDICV010	VL042755.D	1,1,2-Trichloroethane	SAM	7/24/2025 9:03:15 AM	MMDado da	7/24/2025 2:21:15 PM	Peak Integrated by Software incorrectly
VSTDICV010	VL042755.D	m/p-Xylene	SAM	7/24/2025 9:03:15 AM	MMDado da	7/24/2025 2:21:15 PM	Peak Integrated by Software incorrectly

Sequence	VL072425	Instrument	MSVOA_I
----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC010	VL042757.D	m/p-Xylene	SAM	7/25/2025 7:40:22 AM	MMDado da	7/25/2025 4:36:25 PM	Peak Integrated by Software incorrectly
VL0724AB S01	VL042759.D	m/p-Xylene	SAM	7/25/2025 8:06:02 AM	MMDado da	7/25/2025 4:36:24 PM	Peak Integrated by Software incorrectly
Q2671-04DL	VL042760.D	1,1,1-Trichloroethane	SAM	7/25/2025 8:07:19 AM	MMDado da	7/25/2025 4:36:25 PM	Peak Integrated by Software incorrectly
Q2671-04DL	VL042760.D	Carbon Tetrachloride	SAM	7/25/2025 8:07:19 AM	MMDado da	7/25/2025 4:36:25 PM	Peak Integrated by Software incorrectly

Q2671-05DL	VL04276 1.D	Trichloroethene	SAM	7/25/20 25 8:06:07 AM	MMDado da	7/25/20 25 4:36:28 PM	Peak Integrated by Software incorrectly
Q2671-04	VL04276 2.D	4-Methyl-2-Pentanone	SAM	7/25/20 25 4:38:57 PM	MMDado da	7/25/20 25 4:36:30 PM	Peak Integrated by Software incorrectly
Q2671-04	VL04276 2.D	Bromodichloromethane	SAM	7/25/20 25 4:38:57 PM	MMDado da	7/25/20 25 4:36:30 PM	Peak Integrated by Software incorrectly
Q2671-04	VL04276 2.D	Heptane	SAM	7/25/20 25 4:38:57 PM	MMDado da	7/25/20 25 4:36:30 PM	Peak Integrated by Software incorrectly
Q2671-04	VL04276 2.D	Isopropyl Alcohol	SAM	7/25/20 25 4:38:57 PM	MMDado da	7/25/20 25 4:36:30 PM	Peak Integrated by Software incorrectly
Q2671-04	VL04276 2.D	m/p-Xylene	SAM	7/25/20 25 4:38:57 PM	MMDado da	7/25/20 25 4:36:30 PM	Peak Integrated by Software incorrectly
Q2671-04	VL04276 2.D	o-Xylene	SAM	7/25/20 25 4:38:57 PM	MMDado da	7/25/20 25 4:36:30 PM	Peak Integrated by Software incorrectly

Q2671-04	VL04276 2.D	Toluene	SAM	7/25/20 25 4:38:57 PM	MMdado da	7/25/20 25 4:36:30 PM	Peak Integrat ed by Softwar e incorrec tly
Q2671-05	VL04276 3.D	1,1,1- Trichloroethane	sam	7/26/20 25 12:16:1 8 AM	MMdado da	7/26/20 25 12:17:1 7 AM	Peak Integrat ed by Softwar e incorrec tly
Q2671-05	VL04276 3.D	Carbon Tetrachloride	sam	7/26/20 25 12:16:1 8 AM	MMdado da	7/26/20 25 12:17:1 7 AM	Peak Integrat ed by Softwar e incorrec tly
Q2671-05	VL04276 3.D	Heptane	sam	7/26/20 25 12:16:1 8 AM	MMdado da	7/26/20 25 12:17:1 7 AM	Peak Integrat ed by Softwar e incorrec tly
Q2671-05	VL04276 3.D	m/p-Xylene	sam	7/26/20 25 12:16:1 8 AM	MMdado da	7/26/20 25 12:17:1 7 AM	Peak Integrat ed by Softwar e incorrec tly
Q2671-05	VL04276 3.D	Tetrahydrofuran	sam	7/26/20 25 12:16:1 8 AM	MMdado da	7/26/20 25 12:17:1 7 AM	Peak Integrat ed by Softwar e incorrec tly
Q2671-05	VL04276 3.D	Toluene	sam	7/26/20 25 12:16:1 8 AM	MMdado da	7/26/20 25 12:17:1 7 AM	Peak Integrat ed by Softwar e incorrec tly

Q2671-05	VL04276 3.D	Trichloroethene	sam	7/26/20 25 12:16:1 8 AM	MMdado da	7/26/20 25 12:17:1 7 AM	Peak Integrat ed by Softwar e incorrec tly
Q2671-01	VL04276 4.D	Benzene	sam	7/26/20 25 12:16:2 3 AM	MMdado da	7/26/20 25 12:17:2 3 AM	Peak Integrat ed by Softwar e incorrec tly
Q2671-01	VL04276 4.D	Carbon Tetrachloride	sam	7/26/20 25 12:16:2 3 AM	MMdado da	7/26/20 25 12:17:2 3 AM	Peak Integrat ed by Softwar e incorrec tly
Q2671-01	VL04276 4.D	Chlorodifluorometh ane	sam	7/26/20 25 12:16:2 3 AM	MMdado da	7/26/20 25 12:17:2 3 AM	Peak Integrat ed by Softwar e incorrec tly
Q2671-01	VL04276 4.D	Heptane	sam	7/26/20 25 12:16:2 3 AM	MMdado da	7/26/20 25 12:17:2 3 AM	Peak Integrat ed by Softwar e incorrec tly
Q2671-01	VL04276 4.D	Isopropyl Alcohol	sam	7/26/20 25 12:16:2 3 AM	MMdado da	7/26/20 25 12:17:2 3 AM	Peak Integrat ed by Softwar e incorrec tly
Q2671-01	VL04276 4.D	m/p-Xylene	sam	7/26/20 25 12:16:2 3 AM	MMdado da	7/26/20 25 12:17:2 3 AM	Peak Integrat ed by Softwar e incorrec tly

Q2671-01	VL04276 4.D	Toluene	sam	7/26/20 25 12:16:2 3 AM	MMdado da	7/26/20 25 12:17:2 3 AM	Peak Integrat ed by Softwar e incorrec tly
Q2671-01	VL04276 4.D	Trichlorofluoromet hane	sam	7/26/20 25 12:16:2 3 AM	MMdado da	7/26/20 25 12:17:2 3 AM	Peak Integrat ed by Softwar e incorrec tly
Q2671- 01DUP	VL04276 5.D	Benzene	sam	7/26/20 25 12:16:3 0 AM	MMdado da	7/26/20 25 12:17:2 7 AM	Peak Integrat ed by Softwar e incorrec tly
Q2671- 01DUP	VL04276 5.D	Carbon Tetrachloride	sam	7/26/20 25 12:16:3 0 AM	MMdado da	7/26/20 25 12:17:2 7 AM	Peak Integrat ed by Softwar e incorrec tly
Q2671- 01DUP	VL04276 5.D	Chlorodifluorometh ane	sam	7/26/20 25 12:16:3 0 AM	MMdado da	7/26/20 25 12:17:2 7 AM	Peak Integrat ed by Softwar e incorrec tly
Q2671- 01DUP	VL04276 5.D	Dichlorodifluorome thane	sam	7/26/20 25 12:16:3 0 AM	MMdado da	7/26/20 25 12:17:2 7 AM	Peak Integrat ed by Softwar e incorrec tly
Q2671- 01DUP	VL04276 5.D	Heptane	sam	7/26/20 25 12:16:3 0 AM	MMdado da	7/26/20 25 12:17:2 7 AM	Peak Integrat ed by Softwar e incorrec tly

Q2671-01DUP	VL04276 5.D	Isopropyl Alcohol	sam	7/26/20 25 12:16:30 AM	MMdado da	7/26/20 25 12:17:27 AM	Peak Integrated by Software incorrectly
Q2671-01DUP	VL04276 5.D	m/p-Xylene	sam	7/26/20 25 12:16:30 AM	MMdado da	7/26/20 25 12:17:27 AM	Peak Integrated by Software incorrectly
Q2671-01DUP	VL04276 5.D	Toluene	sam	7/26/20 25 12:16:30 AM	MMdado da	7/26/20 25 12:17:27 AM	Peak Integrated by Software incorrectly
Q2671-01DUP	VL04276 5.D	Trichlorofluoromethane	sam	7/26/20 25 12:16:30 AM	MMdado da	7/26/20 25 12:17:27 AM	Peak Integrated by Software incorrectly
Q2671-02	VL04276 7.D	Benzene	SAM	7/25/20 25 8:07:29 AM	MMDado da	7/25/20 25 4:36:40 PM	Peak Integrated by Software incorrectly
Q2671-02	VL04276 7.D	Carbon Tetrachloride	SAM	7/25/20 25 8:07:29 AM	MMDado da	7/25/20 25 4:36:40 PM	Peak Integrated by Software incorrectly
Q2671-02	VL04276 7.D	Heptane	SAM	7/25/20 25 8:07:29 AM	MMDado da	7/25/20 25 4:36:40 PM	Peak Integrated by Software incorrectly

Q2671-02	VL04276 7.D	m/p-Xylene	SAM	7/25/20 25 8:07:29 AM	MMDado da	7/25/20 25 4:36:40 PM	Peak Integrat ed by Softwar e incorrec tly
Q2671-02	VL04276 7.D	Propene	SAM	7/25/20 25 8:07:29 AM	MMDado da	7/25/20 25 4:36:40 PM	Peak Integrat ed by Softwar e incorrec tly
Q2671-02	VL04276 7.D	Tetrachloroethene	SAM	7/25/20 25 8:07:29 AM	MMDado da	7/25/20 25 4:36:40 PM	Peak Integrat ed by Softwar e incorrec tly
Q2671- 02DL	VL04276 8.D	Propene	SAM	7/25/20 25 8:06:27 AM	MMDado da	7/25/20 25 4:36:42 PM	Peak Integrat ed by Softwar e incorrec tly
Q2671-03	VL04276 9.D	Benzene	sam	7/26/20 25 12:16:3 5 AM	MMdado da	7/26/20 25 12:17:3 1 AM	Peak Integrat ed by Softwar e incorrec tly
Q2671-03	VL04276 9.D	Carbon Tetrachloride	sam	7/26/20 25 12:16:3 5 AM	MMdado da	7/26/20 25 12:17:3 1 AM	Peak Integrat ed by Softwar e incorrec tly
Q2671-03	VL04276 9.D	Heptane	sam	7/26/20 25 12:16:3 5 AM	MMdado da	7/26/20 25 12:17:3 1 AM	Peak Integrat ed by Softwar e incorrec tly

Q2671-03	VL04276 9.D	m/p-Xylene	sam	7/26/20 25 12:16:3 5 AM	MMdado da	7/26/20 25 12:17:3 1 AM	Peak Integrat ed by Softwar e incorrec tly
Q2671-03	VL04276 9.D	Methylene Chloride	sam	7/26/20 25 12:16:3 5 AM	MMdado da	7/26/20 25 12:17:3 1 AM	Peak Integrat ed by Softwar e incorrec tly
Q2671-03	VL04276 9.D	Propene	sam	7/26/20 25 12:16:3 5 AM	MMdado da	7/26/20 25 12:17:3 1 AM	Peak Integrat ed by Softwar e incorrec tly
Q2671-03DL	VL04277 0.D	1,1,1-Trichloroethane	SAM	7/25/20 25 8:06:22 AM	MMDado da	7/25/20 25 4:36:46 PM	Peak Integrat ed by Softwar e incorrec tly
Q2671-03DL	VL04277 0.D	1,2-Dichloroethane	SAM	7/25/20 25 8:06:22 AM	MMDado da	7/25/20 25 4:36:46 PM	Peak Integrat ed by Softwar e incorrec tly
Q2671-03DL	VL04277 0.D	Propene	SAM	7/25/20 25 8:06:22 AM	MMDado da	7/25/20 25 4:36:46 PM	Peak Integrat ed by Softwar e incorrec tly
Q2671-03DL	VL04277 0.D	Tetrachloroethene	SAM	7/25/20 25 8:06:22 AM	MMDado da	7/25/20 25 4:36:46 PM	Peak Integrat ed by Softwar e incorrec tly

Q2671-03DL	VL042770.D	Toluene	SAM	7/25/2025 8:06:22 AM	MMDado da	7/25/2025 4:36:46 PM	Peak Integrated by Software incorrectly
VSTDCCC010	VL042771.D	m/p-Xylene	SAM	7/25/2025 8:06:17 AM	MMDado da	7/25/2025 4:36:51 PM	Peak Integrated by Software incorrectly

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q2671

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 07/31/2025