

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: RTPE01
 Lab Code: ACE SDG No.: Q2671
 Instrument ID: MSVOA_L Calibration Date/Time: 07/24/2025 08:23
 Lab File ID: VL042757.D Init. Calib. Date(s): 07/23/2025 07/23/2025
 Heated Purge: (Y/N) N Init. Calib. Time(s): 08:58 13:00
 GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
1,2,4-Trichlorobenzene	10.000	10.16	-1.569996	20
Naphthalene	10.000	10.19	-1.850004	20