

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL072325

Review By	Semsettin Yesilyurt	Review On	7/24/2025 9:04:14 AM
Supervise By	Maresh Dadoda	Supervise On	7/24/2025 2:21:30 PM
SubDirectory	VL072325	HP Acquire Method	TO15AIR.M
		HP Processing Method	VL072325AIR.M

STD. NAME	STD REF.#
Tune/Reschk	AP2645
Initial Calibration Stds	AP2641,AP2642,AP2643
CCC	AP2641
Internal Standard/PEM	AP2645
ICV/I.BLK	AP2644
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VL042746.D	23 Jul 2025 08:00		SY/MD	Ok
2	VSTDICCC010	VSTDICCC010	VL042747.D	23 Jul 2025 08:58		SY/MD	Ok,M
3	VSTDICC002	VSTDICC002	VL042748.D	23 Jul 2025 09:34		SY/MD	Ok,M
4	VSTDICC001	VSTDICC001	VL042749.D	23 Jul 2025 10:08	Comp. #13,79,80,81 are on Linear Regression	SY/MD	Ok,M
5	VSTDICC0.5	VSTDICC0.5	VL042750.D	23 Jul 2025 10:42		SY/MD	Ok,M
6	VSTDICC0.1	VSTDICC0.1	VL042751.D	23 Jul 2025 11:16	Not used	SY/MD	Not Ok
7	VSTDICC0.03	VSTDICC0.03	VL042752.D	23 Jul 2025 11:50		SY/MD	Ok,M
8	VSTDICC015	VSTDICC015	VL042753.D	23 Jul 2025 12:26	%D failed for Comp. #80 in 0.5 and 2 ppb	SY/MD	Ok,M
9	VSTDICC0.1	VSTDICC0.1	VL042754.D	23 Jul 2025 13:00	%D failed for Comp. #79 in 0.1 ppb	SY/MD	Ok,M
10	VSTDICV010	ICVVVL072325	VL042755.D	23 Jul 2025 14:33		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL072425

Review By	Semsettin Yesilyurt	Review On	7/25/2025 8:10:00 AM
Supervise By	Maresh Dadoda	Supervise On	7/25/2025 4:36:51 PM
SubDirectory	VL072425	HP Acquire Method	TO15AIR.M
		HP Processing Method	VL072325AIR.M

STD. NAME	STD REF.#
Tune/Reschk	AP2645
Initial Calibration Stds	AP2641,AP2642,AP2643
CCC	AP2641
Internal Standard/PEM	AP2645
ICV/I.BLK	AP2644
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VL042756.D	24 Jul 2025 07:39		SY/MD	Ok
2	VSTDCCC010	VSTDCCC010	VL042757.D	24 Jul 2025 08:23		SY/MD	Ok,M
3	VL0724ABL01	VL0724ABL01	VL042758.D	24 Jul 2025 10:12		SY/MD	Ok
4	VL0724ABS01	VL0724ABS01	VL042759.D	24 Jul 2025 10:57		SY/MD	Ok,M
5	Q2671-04DL	118-SG2DL	VL042760.D	24 Jul 2025 11:39		SY/MD	Ok,M
6	Q2671-05DL	118-SG1DL	VL042761.D	24 Jul 2025 12:14		SY/MD	Ok,M
7	Q2671-04	118-SG2	VL042762.D	24 Jul 2025 13:17	Need 10X	SY/MD	Dilution
8	Q2671-05	118-SG1	VL042763.D	24 Jul 2025 13:52	Need 10X	SY/MD	Dilution
9	Q2671-01	118-AAI	VL042764.D	24 Jul 2025 14:28		SY/MD	Ok,M
10	Q2671-01DUP	118-AAIDUP	VL042765.D	24 Jul 2025 15:05		SY/MD	Ok,M
11	Q2671-01DL	118-AAIDL	VL042766.D	24 Jul 2025 15:40	Not Required	SY/MD	Not Ok
12	Q2671-02	118-IA-1	VL042767.D	24 Jul 2025 16:17	Need 10X	SY/MD	Dilution
13	Q2671-02DL	118-IA-1DL	VL042768.D	24 Jul 2025 16:53		SY/MD	Ok,M
14	Q2671-03	118-IA-2	VL042769.D	24 Jul 2025 17:29	Need 10X	SY/MD	Dilution
15	Q2671-03DL	118-IA-2DL	VL042770.D	24 Jul 2025 18:05		SY/MD	Ok,M
16	VSTDCCC010	VSTDCCC010EC	VL042771.D	24 Jul 2025 19:08		SY/MD	Ok,M

M : Manual Integration