



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

**Core Environmental Consultants and Services, Inc.**

**Project Name: NYPA**

**Project # N/A**

**Order ID # Q2681**

**Test Name: Ignitability,pH**

**A. Number of Samples and Date of Receipt:**

1 Solid sample was received on 07/23/2025.

**B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: Ignitability,pH. This data package contains results for Ignitability,pH.

**C. Analytical Techniques:**

The analysis of Ignitability was based on method 1030 and The analysis of pH was based on method 9045D.

**D. QA/ QC Samples:**

The Holding Times were met for all samples except for NYPA-POUCH-SPENT-CARBON of pH as sample was receive out of holding time.

The Duplicate analysis met criteria for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

**E. Additional Comments:**

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_