

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136593

Slope : 98.3

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.02	07/24/2025	11:50
2	CAL2	1	Water	NA	NA	20.2	7.01	07/24/2025	11:51
3	CAL3	1	Water	NA	NA	20.3	10.02	07/24/2025	11:55
4	ICV	1	Water	NA	NA	20.3	7.02	07/24/2025	11:57
5	CCV1	1	Water	NA	NA	20.3	2.01	07/24/2025	11:59
6	Q2681-01	1	Solid	20.02	20	22.6	5.46	07/24/2025	12:15
7	Q2681-01DUP	1	Solid	20.04	20	22.8	5.48	07/24/2025	12:19
8	CCV2	1	Water	NA	NA	20.3	12.02	07/24/2025	12:20

WORKLIST(Hardcopy Internal Chain)

17131593

WorkList Name : ph s q2681

WorkList ID : 190918

Department : Wet-Chemistry

Date : 07-24-2025 11:31:45

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2681-01	NYPA-POUCH-SPENT-CARBO	Solid	pH	Cool 4 deg C	CORE02	D21	07/23/2025	9045D

Date/Time

07/24/25 11:45

Raw Sample Received by:

gc (wp)

Raw Sample Relinquished by:

gc (sum)

Date/Time

07/24/25

Raw Sample Received by:

gc (wp)

Raw Sample Relinquished by:

gc (sum)