



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Core Environmental Consultants and Services, Inc.

SDG No.: Q2681

Contract: CORE02

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.23	0.2	114	70 - 130	CV	07/28/2025	11:57	LB136633
CRI01	Arsenic	20.1	20.0	101	65 - 135	P	07/28/2025	12:07	LB136640
	Barium	104	100	104	65 - 135	P	07/28/2025	12:07	LB136640
	Cadmium	5.58	6.0	93	65 - 135	P	07/28/2025	12:07	LB136640
	Chromium	10.7	10.0	107	65 - 135	P	07/28/2025	12:07	LB136640
	Lead	10.9	12.0	90	65 - 135	P	07/28/2025	12:07	LB136640
	Selenium	19.5	20.0	98	65 - 135	P	07/28/2025	12:07	LB136640
	Silver	10.6	10.0	106	65 - 135	P	07/28/2025	12:07	LB136640