



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Core Environmental Consultants and Services, Inc.

SDG No.: Q2681

Contract: CORE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB40	Mercury	0.076	+/-0.2	U	0.20	CV	07/28/2025	11:48	LB136633

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Core Environmental Consultants and Services, Inc.

SDG No.: Q2681

Contract: CORE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB51	Mercury	0.076	+/-0.2	U	0.20	CV	07/28/2025	11:52	LB136633
CCB52	Mercury	0.076	+/-0.2	U	0.20	CV	07/28/2025	12:25	LB136633
CCB53	Mercury	0.076	+/-0.2	U	0.20	CV	07/28/2025	12:52	LB136633
CCB54	Mercury	0.076	+/-0.2	U	0.20	CV	07/28/2025	13:13	LB136633
CCB55	Mercury	0.076	+/-0.2	U	0.20	CV	07/28/2025	13:27	LB136633

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Core Environmental Consultants and Services, Inc.

SDG No.: Q2681

Contract: CORE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	07/28/2025	11:57	LB136640
	Barium	14.6	+/-50	U	100	P	07/28/2025	11:57	LB136640
	Cadmium	0.50	+/-3	U	6.00	P	07/28/2025	11:57	LB136640
	Chromium	2.12	+/-5	U	10.0	P	07/28/2025	11:57	LB136640
	Lead	2.30	+/-6	U	12.0	P	07/28/2025	11:57	LB136640
	Selenium	9.64	+/-10	U	20.0	P	07/28/2025	11:57	LB136640
	Silver	1.62	+/-5	U	10.0	P	07/28/2025	11:57	LB136640

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Core Environmental Consultants and Services, Inc.

SDG No.: Q2681

Contract: CORE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	07/28/2025	12:33	LB136640
	Barium	14.6	+/-50	U	100	P	07/28/2025	12:33	LB136640
	Cadmium	0.50	+/-3	U	6.00	P	07/28/2025	12:33	LB136640
	Chromium	2.12	+/-5	U	10.0	P	07/28/2025	12:33	LB136640
	Lead	2.30	+/-6	U	12.0	P	07/28/2025	12:33	LB136640
	Selenium	9.64	+/-10	U	20.0	P	07/28/2025	12:33	LB136640
	Silver	1.62	+/-5	U	10.0	P	07/28/2025	12:33	LB136640
CCB02	Arsenic	5.12	+/-10	U	20.0	P	07/28/2025	13:31	LB136640
	Barium	14.6	+/-50	U	100	P	07/28/2025	13:31	LB136640
	Cadmium	0.50	+/-3	U	6.00	P	07/28/2025	13:31	LB136640
	Chromium	2.12	+/-5	U	10.0	P	07/28/2025	13:31	LB136640
	Lead	2.30	+/-6	U	12.0	P	07/28/2025	13:31	LB136640
	Selenium	9.64	+/-10	U	20.0	P	07/28/2025	13:31	LB136640
	Silver	1.62	+/-5	U	10.0	P	07/28/2025	13:31	LB136640
CCB03	Arsenic	5.12	+/-10	U	20.0	P	07/28/2025	14:42	LB136640
	Barium	14.6	+/-50	U	100	P	07/28/2025	14:42	LB136640
	Cadmium	0.50	+/-3	U	6.00	P	07/28/2025	14:42	LB136640
	Chromium	2.12	+/-5	U	10.0	P	07/28/2025	14:42	LB136640
	Lead	2.30	+/-6	U	12.0	P	07/28/2025	14:42	LB136640
	Selenium	9.64	+/-10	U	20.0	P	07/28/2025	14:42	LB136640
	Silver	1.62	+/-5	U	10.0	P	07/28/2025	14:42	LB136640
CCB04	Arsenic	5.12	+/-10	U	20.0	P	07/28/2025	16:08	LB136640
	Barium	14.6	+/-50	U	100	P	07/28/2025	16:08	LB136640
	Cadmium	0.50	+/-3	U	6.00	P	07/28/2025	16:08	LB136640
	Chromium	2.12	+/-5	U	10.0	P	07/28/2025	16:08	LB136640
	Lead	2.30	+/-6	U	12.0	P	07/28/2025	16:08	LB136640
	Selenium	9.64	+/-10	U	20.0	P	07/28/2025	16:08	LB136640
	Silver	1.62	+/-5	U	10.0	P	07/28/2025	16:08	LB136640
CCB05	Arsenic	5.12	+/-10	U	20.0	P	07/28/2025	17:02	LB136640
	Barium	14.6	+/-50	U	100	P	07/28/2025	17:02	LB136640
	Cadmium	0.50	+/-3	U	6.00	P	07/28/2025	17:02	LB136640
	Chromium	2.12	+/-5	U	10.0	P	07/28/2025	17:02	LB136640
	Lead	2.30	+/-6	U	12.0	P	07/28/2025	17:02	LB136640
	Selenium	9.64	+/-10	U	20.0	P	07/28/2025	17:02	LB136640
	Silver	1.62	+/-5	U	10.0	P	07/28/2025	17:02	LB136640